



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>BARRY VARRATTA</i>		
Location <i>free</i>	Intoxilyzer® Serial Number <i>80-006688</i>	
Gas Lot Number <i>0221080 A1</i>	Gas Cylinder Number <i>34</i>	Gas Expiration Date <i>3-5-2023</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.081* AC
  - B. *0.081* AC
  - C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Barry Varratta</i>	Date <i>2.24.2021</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Ehl</i>	Date <i>3/2/21</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer . . . Alcohol Analyzer  
North Dakota Model 8000 . . . SN 80-006688  
Location = STEE . . . 8164.14.00 09/16  
02/24/2021 . . . 16:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:26
02 Std. Gas	0.081	16:26
03 Room Air	0.000	16:27
04 Std. Gas	0.081	16:27
05 Room Air	0.000	16:28
06 Std. Gas	0.080	16:28
07 Room Air	0.000	16:28

Lot No = 02621080A1  
Cyl No = 34  
Exp Date = 03/05/2023  
County = 38                      Oper No. = 039909

  
\_\_\_\_\_  
Operator Signature  
BARRY VANNATTA

Remarks:

Form 106-I8000