




INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number <u>80-006512</u>	Instrument Location <u>Carrington PD (CARR)</u>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature  #1622	Date <u>5/21/2021</u>
--	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date <u>5/24/2021</u>
--	------------------------------------

I certify that this document is a true and correct duplicate of the original.

Name Hutton #1622
 Date 5/21/2021
 Police Department
 Carrington, N.D.

Intoxilyzer Test Record and Checklist
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
 North Dakota Model 8000 SN 80-006512
 Location = CARR 8164.14.00 09/16
 05/21/2021 20:00

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=-|
 ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=-|
 ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
 Start Test Sequence: DABACABA
 Display Prelim Rslt? Yes
 Display Third Digit? Yes
 Inhib Printer(Y/N)? No
 Display Volume? No
 Disable On Memfull? Yes
 # of Print Copies? 1
 Select Std (D/W/I)? Dry
 Standard Value? 0.080
 Standard Lot #? 24119080A1
 Standard Cyl #? 13
 Standard Expiration? 11/05/2021
 Oper No? 132903

Flow Cal. Date: 08/20/2015
 Slope 680
 Intercept -660856


IR Calibration Date: 08/21/2015

	3um	9um
--	-----	-----

0th Coef(*100):	-23051	-19261
1st Coef(*100):	266836	137135
2nd Coef(*100):	3974	1466
H2O adj(mg/l*10k):	785	487

***** Printer Test End *****

I certify that this document
 is a true and correct duplicate
 of the original.
 Name Hutton #1622
 Date 5/21/2021
 Police Department
 Carrington, N.D.



 Operator Signature
 RYAN HUTTON

Remarks: Print Test


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/21/2021 20:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:03
02 Std. Gas	0.080	20:03
03 Room Air	0.000	20:04
04 Std. Gas	0.080	20:04
05 Room Air	0.000	20:05
06 Std. Gas	0.080	20:05
07 Room Air	0.000	20:06

Lot No = 24119080A1
Cyl No = 13
Exp Date = 11/05/2021
County = 16 Oper No. = 132903

 #1622
Operator Signature
RYAN HUTTON

Remarks: ACA Test

Form 106-I8000

I certify that this document
is a true and correct duplicate
of the original.
Name Hutton #1622
Date 5/21/2021
Police Department
Carrington, N.D.

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/21/2021 20:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:07
02 Std. Gas	0.081	20:07
03 Room Air	0.000	20:08
04 Std. Gas	RFI*	20:08
05 Room Air	RFI*	20:08

*Invalid Test
Inhibited - RFI

Lot No = 24119080A1
Cyl No = 13
Exp Date = 11/05/2021
County = 16 Oper No. = 132903


#1622
Operator Signature
RYAN HUTTON

Remarks: RFI Test

Form 106-I8000

I certify that this document
is a true and correct duplicate
of the original.
Name Hutton, #1622
Date 5/21/2021
Police Department
Carrington, N.D.