

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006688

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 996 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP3064 Lot #: 20070 Exp. Date: 2.13.22
- b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: MP5317 Lot #: 202012A Exp. Date: 12.8.22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: ICS 6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 24119080A1 Cyl #: 31 Exp. Date: 11.5.21
Test 1 0.082 Test 4 0.082 Test 7 0.082
Test 2 0.082 Test 5 0.083 Test 8 0.083
Test 3 0.082 Test 6 0.083 Test 9 0.082
Average 0.082

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED
EXTERNAL BATTERY

Instrument is acceptable to be used in the field.

Charles E. Eder
Breath Analyst Signature

5/26/2021
Date

Kali Z. Hoot
Reviewed by

5.26.21
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/26/2021 13:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 34
Standard Expiration? 03/05/2023
Oper No? 666666

Flow Cal. Date: 09/07/2016
Slope 698
Intercept -642832

IR Calibration Date: 09/07/2016

	3um	9um
0th Coef(*100):	-10822	-23152
1st Coef(*100):	262468	134492
2nd Coef(*100):	3796	1436
H2O adj(mg/l*10k):	665	568

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/26/2021 13:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:46
02 Std. Sol.	0.018	13:46
03 Room Air	0.000	13:47
04 Std. Sol.	0.018	13:48
05 Room Air	0.000	13:48
06 Std. Sol.	0.019	13:49
07 Room Air	0.000	13:49

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 20070

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/26/2021 13:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Sol.	0.298	13:52
03 Room Air	0.000	13:53
04 Std. Sol.	0.301	13:53
05 Room Air	0.000	13:54
06 Std. Sol.	0.300	13:54
07 Room Air	0.000	13:55

08 Sim Temp = 34.0°C

Simul Ser No = MP5317
Std Sol No = 202012A
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/26/2021 13:57

Test	AC	Time
01 Room Air	0.000	13:57
02 *Subject Test	INT*	13:58
03 Room Air	0.000	13:58

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 24119080A1
Cyl No = 31
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/26/2021 14:01

Test	AC	Time
01 Diagnostic	OK	14:02
02 Room Air	0.000	14:02
03 *Subject Test	RFI*	14:02
04 Room Air	0.000	14:03

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 24119080A1
Cyl No = 31
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/26/2021 14:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:05
02 Std. Gas	0.082	14:05
03 Room Air	0.000	14:06
04 Std. Gas	0.082	14:06
05 Room Air	0.000	14:06
06 Std. Gas	0.082	14:07
07 Room Air	0.000	14:07

Lot No = 24119080A1
Cyl No = 31
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
CEE 0.082 0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
05/26/2021 14:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:08
02 Std. Gas	0.082	14:09
03 Room Air	0.000	14:09
04 Std. Gas	0.083	14:09
05 Room Air	0.000	14:10
06 Std. Gas	0.083	14:10
07 Room Air	0.000	14:11

Lot No = 24119080A1
Cyl No = 31
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

