

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006686

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 916 psi Regulator: 900 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.030 AC
Sim. Ser #: MP5289 Lot #: 20911E Exp. Date: 11.19.21
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3062 Lot #: 202103E Exp. Date: 3.24.23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: ICS 6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 24119080A1 Cyl #: 5 Exp. Date: 11.5.21

Test 1	<u>0.081</u>	Test 4	<u>0.081</u>	Test 7	<u>0.080</u>
Test 2	<u>0.081</u>	Test 5	<u>0.081</u>	Test 8	<u>0.080</u>
Test 3	<u>0.081</u>	Test 6	<u>0.080</u>	Test 9	<u>0.081</u>
Average	<u>0.081</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED EXTERNAL BATTERY

Instrument is acceptable to be used in the field.

Breath Analyst Signature

Reviewed by

Date

Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = TOXL 8164.14.00 09/16
06/16/2021 13:26

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%\$^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%\$^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 6
Standard Expiration? 03/05/2023
Oper No? 666666

Flow Cal. Date: 09/07/2016
Slope 770
Intercept -749410

IR Calibration Date: 09/06/2016
 3um 9um

0th Coef(*100): -14519 -24061
1st Coef(*100): 252211 131592
2nd Coef(*100): 3527 1490
H2O adj(mg/l*10k): 522 594

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = TOXL 8164.14.00 09/16
06/16/2021 13:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:30
02 Std. Sol.	0.028	13:31
03 Room Air	0.000	13:32
04 Std. Sol.	0.028	13:32
05 Room Air	0.000	13:33
06 Std. Sol.	0.029	13:34
07 Room Air	0.000	13:34


08 Sim Temp = 34.0°C

Simul Ser No = MP5289

Std Sol No = 201911E

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC

0.030 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = TOXL 8164.14.00 09/16
06/16/2021 13:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:36
02 Std. Sol.	0.395	13:37
03 Room Air	0.000	13:38
04 Std. Sol.	0.394	13:38
05 Room Air	0.000	13:39
06 Std. Sol.	0.395	13:40
07 Room Air	0.000	13:40

08 Sim Temp = 34.0°C

Simul Ser No = MP3062
Std Sol No = 202103E
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
 0.400 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = TOXL 8164.14.00 09/16
06/16/2021 13:42

Test	AC	Time
01 Room Air	0.000	13:43
02 *Subject Test	INT*	13:43
03 Room Air	0.000	13:44

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 24119080A1
Cyl No = 5
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = TOXL 8164.14.00 09/16
06/16/2021 13:45

Test	AC	Time
01 Diagnostic	OK	13:45
02 Room Air	0.000	13:46
03 *Subject Test	RFI*	13:46
04 Room Air	0.000	13:47

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 24119080A1
Cyl No = 5
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = TOXL 8164.14.00 09/16
06/16/2021 13:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:48
02 Std. Gas	0.081	13:48
03 Room Air	0.000	13:49
04 Std. Gas	0.081	13:49
05 Room Air	0.000	13:50
06 Std. Gas	0.081	13:50
07 Room Air	0.000	13:51

Lot No = 24119080A1
Cyl No = 5
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = TOXL 8164.14.00 09/16
06/16/2021 13:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Gas	0.081	13:52
03 Room Air	0.000	13:52
04 Std. Gas	0.081	13:53
05 Room Air	0.000	13:53
06 Std. Gas	0.080	13:53
07 Room Air	0.000	13:54

Lot No = 24119080A1
Cyl No = 5
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = TOXL 8164.14.00 09/16
06/16/2021 13:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:55
02 Std. Gas	0.080	13:55
03 Room Air	0.000	13:56
04 Std. Gas	0.080	13:56
05 Room Air	0.000	13:57
06 Std. Gas	0.081	13:57
07 Room Air	0.000	13:57

Lot No = 24119080A1
Cyl No = 5
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000