

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006679 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 815 psi Regulator: 800 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.015 AC
Sim. Ser #: MP3067 Lot #: 201911C Exp. Date: 11-19-21
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.250 AC
Sim. Ser #: MP5317 Lot #: 201911B Exp. Date: 11-5-21

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: ICS 6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 24119080A1 Cyl #: 33 Exp. Date: 11-5-21

Test 1	<u>0.081</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.081</u>	Test 8	0.081 CEE 0.081
Test 3	<u>0.081</u>	Test 6	<u>0.082</u>	Test 9	<u>0.081</u>
Average <u>0.081</u>					

C. Remarks/Maintenance: INSPECTION AFTER BEING RETURNED
BY VALLEY CITY P.D. BARNES COUNTY & VALLEY CITY PD.
WILL BE SHARING AN INSTRUMENT AT NEW CORRECTIONS/
LEC FACILITY.

Instrument is acceptable to be used in the field.

Breath Analyst Signature

Reviewed by

Date

Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = TOXL 8164.14.00 09/16
10/11/2021 12:59

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 33
Standard Expiration? 11/05/2021
Oper No? 666666

Flow Cal. Date: 09/06/2016
Slope 701
Intercept -682926

IR Calibration Date: 09/06/2016
 3um 9um

0th Coef(*100): -18983 -23165
1st Coef(*100): 280986 134481
2nd Coef(*100): 2443 1290
H2O adj(mg/l*10k): 569 553

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = TOXL 8164.14.00 09/16
10/11/2021 13:39

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:39
02 Std. Sol.	0.016	13:40
03 Room Air	0.000	13:41
04 Std. Sol.	0.016	13:41
05 Room Air	0.000	13:42
06 Std. Sol.	0.015	13:43
07 Room Air	0.000	13:43


08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 201911C

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

LOW AC
0.015 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = TOXL 8164.14.00 09/16
10/11/2021 13:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:45
02 Std. Sol.	0.248	13:46
03 Room Air	0.000	13:47
04 Std. Sol.	0.249	13:48
05 Room Air	0.000	13:48
06 Std. Sol.	0.249	13:49
07 Room Air	0.000	13:49

08 Sim Temp = 34.0°C

Simul Ser No = MP5317

Std Sol No = 210911B

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.250 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = TOXL 8164.14.00 09/16
10/11/2021 14:03

Test	AC	Time
01 Room Air	0.000	14:04
02 *Subject Test	INT*	14:04
03 Room Air	0.000	14:05

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 24119080A1
Cyl No = 33
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = TOXL 8164.14.00 09/16
10/11/2021 13:54

Test	AC	Time
01 Diagnostic	OK	13:55
02 Room Air	0.000	13:56
03 *Subject Test	RFI*	13:56
04 Room Air	0.000	13:56

*Invalid Test
 Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 24119080A1
Cyl No = 33
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = TOXL 8164.14.00 09/16
10/11/2021 14:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:05
02 Std. Gas	0.081	14:06
03 Room Air	0.000	14:06
04 Std. Gas	0.081	14:07
05 Room Air	0.000	14:07
06 Std. Gas	0.081	14:08
07 Room Air	0.000	14:08

Lot No = 24119080A1
Cyl No = 33
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = TOXL 8164.14.00 09/16
10/11/2021 14:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:09
02 Std. Gas	0.081	14:09
03 Room Air	0.000	14:10
04 Std. Gas	0.081	14:10
05 Room Air	0.000	14:11
06 Std. Gas	0.082	14:11
07 Room Air	0.000	14:12

Lot No = 24119080A1
Cyl No = 33
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = TOXL 8164.14.00 09/16
10/11/2021 14:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:13
02 Std. Gas	0.081	14:13
03 Room Air	0.000	14:14
04 Std. Gas	0.081	14:14
05 Room Air	0.000	14:15
06 Std. Gas	0.081	14:15
07 Room Air	0.000	14:16

Lot No = 24119080A1
Cyl No = 33
Exp Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000