



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Eric Braathen</i>			<i>CEE 11.5.21</i>
Location <i>Killdeer PD</i>	<i>CEE 11.5.21</i>	Intoxilyzer® Serial Number <i>0262108 A1 80-006676</i>	
Gas Lot Number <i>5</i>	Gas Cylinder Number <i>005</i>	Gas Expiration Date <i>3/5/23</i>	

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. *0.082* AC  
 B. *0.091* AC  
 C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:  
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).  
 B. ACA Test Record.

Chemical Test Operator Signature <i>Eric Braathen</i>	Date <i>11/4/21</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. ...</i>	Date <i>11/5/2021</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = KDPD      8164.14.00 09/16  
11/04/2021      22:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:05
02 Std. Gas	0.082	22:05
03 Room Air	0.000	22:06
04 Std. Gas	0.081	22:06
05 Room Air	0.000	22:07
06 Std. Gas	0.080	22:07
07 Room Air	0.000	22:08

Lot No = 02621080A1  
Cyl No = 5  
Exp Date = 03/05/2023  
County = 13      Oper No. = 100444



Operator Signature  
ERIC BRAATHEN

Remarks:

Form 106-I8000