

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-006682

Agency being assigned instrument: O.A.G. - CRIME LAB - BREATH

Date 3/10/22

Transferred from:

Name (print) Doreen Rustebekke

Signature [Signature]

Agency VALLEY CITY P.D.

Transferred to:

Name (print) CHARLES E. EDER

Signature [Signature]

Agency O.A.G. - CRIME LAB - BREATH

Reason for transfer/remarks:

- INSTRUMENT DISPLAYING "DISABLED, EEPROM ERROR #016". WILL NOT CLEAR ON RESTART.

- BEING REPLACED BY 80-004943

Or Shipped Via:

USPS UPS FedEx Other

Date Sending Instrument: / / Air Bill No.