



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) MARK PEARSON		
Location EMMONS COUNTY	Intoxilyzer® Serial Number 80-006689	
Gas Lot Number 34921080A1	Gas Cylinder Number 56	Gas Expiration Date 2/5/24

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. **0.078** AC
 B. **0.072** AC
 C. **0.079** AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature	Date 4/4/22
Reviewed By (Crime Laboratory Use Only) Janelle Pertschiller	Date 12 Apr 22

Intoxilyzer Test Record and Checklist
 NDCC Crime Lab. Div., Bismarck, ND 58501
 CMI, Inc. Intoxilyzer Alcohol Analyzer
 North Dakota Model 8090 SN 80-006689
 Location = LINT 8164.14.00 09/15
 04/04/2022 09:12

DRY CAL CHECK

Test	g/c	Time
01 Room Air	0.000	09:13
02 Std. Gas	0.078	09:13
03 Room Air	0.000	09:14
04 Std. Gas	0.077	09:14
05 Room Air	0.000	09:15
06 Std. Gas	0.079	09:15
07 Room Air	0.000	09:16

LOT: NO = 34921080A1
 Cyl No = 56
 Exp Date = 02/05/2024
 County = 15
 Operator Signature
 MARK PEARSON

Form 106-18000