



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Thomas, Patrick</i>		
Location <i>WFPD</i>	Intoxilyzer® Serial Number <i>80-006684</i>	
Gas Lot Number <i>26021080A1</i>	Gas Cylinder Number <i>2</i>	Gas Expiration Date <i>10/05/2023</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.081* AC
  - B. *0.082* AC
  - C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Patrick Thomas</i>	Date <i>04/04/2022</i>
Reviewed By (Crime Laboratory Use Only) <i>Jenelle Porscheller</i>	Date <i>12 Apr 22</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-006684  
Location = WFPD                              8164.14.00 09/16  
04/04/2022    22:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:11
02 Std. Gas	0.081	22:11
03 Room Air	0.000	22:12
04 Std. Gas	0.082	22:12
05 Room Air	0.000	22:13
06 Std. Gas	0.081	22:13
07 Room Air	0.000	22:14

Lot No = 26021080A1  
Cyl No = 2  
Exp Date = 10/05/2023  
County = 09    Oper No. = 130056

 #712

Operator Signature  
PATRICK THOMAS

Remarks: *Installation test*