



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-006684	Instrument Location WFPD
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): None (Return from Annual Inspection)
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Feld + [Signature]</i>	Date 04/04/2022
--	--------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Janelle Pertscheller</i>	Certified Date 12 Apr 22
--	-----------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = WFPD 8164.14.00 09/16
04/04/2022 22:07

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 2
Standard Expiration? 10/05/2023
Oper No? 130056

Flow Cal. Date: 09/06/2016
Slope 709
Intercept -725441

IR Calibration Date: 09/06/2016
 3um 9um

0th Coef (*100): -16013 -23467
1st Coef (*100): 256431 135731
2nd Coef (*100): 2904 1268
H2O adj (mg/l*10k): 553 514

***** Printer Test End *****



Operator Signature
PATRICK THOMAS

Remarks: Installation test


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = WFPD 8164.14.00 09/16
04/04/2022 22:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:11
02 Std. Gas	0.081	22:11
03 Room Air	0.000	22:12
04 Std. Gas	0.082	22:12
05 Room Air	0.000	22:13
06 Std. Gas	0.081	22:13
07 Room Air	0.000	22:14

Lot No = 26021080A1
Cyl No = 2
Exp Date = 10/05/2023
County = 09 Oper No. = 130056

 #712

Operator Signature
PATRICK THOMAS

Remarks: *Installation test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = WFPD 8164.14.00 09/16
04/04/2022 22:20

Test	AC	Time
01 Room Air	RFI*	22:22
02 Room Air	0.000	22:22

*Invalid Test
Inhibited - RFI

Sub Name = TEST, INSTALLATION NA
Sub DOB = 04/04/2000
Sub Sex = Male Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 26021080A1
Cyl No = 2
Expiration Date = 10/05/2023
County = 09 Oper No. = 130056

 # 712

Operator Signature
PATRICK THOMAS

Remarks: Installation test