

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006507

Location: TOXL

A. General Setup and Checks:

- 1. Diagnostic Tests Pass and Instrument in Ready Mode
- 2. Breath tube heated
- 3. Date, time and location code (Level 2,E). Re-set if necessary.
- 4. Print test (Level 1,P). Sign and attach test record.
- 5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 300^{psi} / 302 psi Regulator: 300 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3060 Lot #: 202201A Exp. Date: 01/11/24
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12/8/22

Configure dry gas standard for the following tests (Level 1,S).

- 2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1C56 Exp. Date: NA
 Display reads "Interferent Detect"
- 3. RFI Check (CMS Mode) Display reads "RFI Detect"
- 4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 4 Exp. Date: 3/5/23

Test 1	<u>0.081</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.081</u>	Test 9	<u>0.081</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: Annual inspection

Replaced breath tube o-ring and simulator return o-ring.
Replaced external battery.

Instrument is acceptable to be used in the field.

Janelle Pertschiller
Breath Analyst Signature

6/8/2022
Date

[Signature]
Reviewed by

06/3/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = GFSO 8164.14.00 09/16
06/08/2022 16:05

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^ & * () _ + ?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^ & * () _ + ?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer (Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 16
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 08/20/2015
Slope 659
Intercept -658315

IR Calibration Date: 08/20/2015
 3um 9um

0th Coef(*100): -9452 -11722
1st Coef(*100): 258808 139588
2nd Coef(*100): 4476 1674
H2O adj (mg/l*10k): 499 376

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = TOXL 8164.14.00 09/16
06/08/2022 16:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:11
02 Std. Sol.	0.020	16:12
03 Room Air	0.000	16:13
04 Std. Sol.	0.020	16:13
05 Room Air	0.000	16:14
06 Std. Sol.	0.020	16:15
07 Room Air	0.000	16:15

08 Sim Temp = 34.0°C

Simul Ser No = MP3060
Std Sol No = 202201A
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low Ac check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = TOXL 8164.14.00 09/16
06/08/2022 16:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:17
02 Std. Sol.	0.300	16:17
03 Room Air	0.000	16:18
04 Std. Sol.	0.303	16:19
05 Room Air	0.000	16:19
06 Std. Sol.	0.304	16:20
07 Room Air	0.000	16:21

08 Sim Temp = 34.0°C

Simul Ser No = MP3064
Std Sol No = 202012A
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Err High Ac check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = TOXL 8164.14.00 09/16
06/08/2022 16:31

Test	AC	Time
01 Room Air	0.000	16:32
02 *Subject Test	INT*	16:32
03 Room Air	0.000	16:33

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 4
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = TOXL 8164.14.00 09/16
06/08/2022 16:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:43
02 Std. Gas	0.081	16:43
03 Room Air	0.000	16:44
04 Std. Gas	0.080	16:44
05 Room Air	0.000	16:45
06 Std. Gas	0.081	16:45
07 Room Air	0.000	16:45

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = TOXL 8164.14.00 09/16
06/08/2022 16:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:48
02 Std. Gas	0.081	16:48
03 Room Air	0.000	16:48
04 Std. Gas	0.080	16:49
05 Room Air	0.000	16:49
06 Std. Gas	0.081	16:50
07 Room Air	0.000	16:50

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128

Janelle Portscheller
Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = TOXL 8164.14.00 09/16
06/08/2022 16:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:51
02 Std. Gas	0.081	16:52
03 Room Air	0.000	16:52
04 Std. Gas	0.081	16:53
05 Room Air	0.000	16:53
06 Std. Gas	0.081	16:53
07 Room Air	0.000	16:54

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check #7-9*