

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006677 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 473 psi Regulator: 560 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3060 Lot #: 202201A Exp. Date: 4/11/24
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12/8/22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1ESG Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 020 Exp. Date: 3/5/2023

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.080</u>	Test 8	<u>0.081</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average <u>0.080</u>					

C. Remarks/Maintenance: Annual Inspection.
Replaced breath tube - o-ring and simulator return o-ring.
Replaced external battery.

Instrument is acceptable to be used in the field.

Janelle Portochillo
Breath Analyst Signature

06/10/22
Date

[Signature]
Reviewed by

06/322
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006677
Location = TOXL 8164.14.00 09/16
06/10/2022 09:34

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 8
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 08/24/2016
Slope 696
Intercept -655873

IR Calibration Date: 08/24/2016

	3um	9um
0th Coef(*100):	-12424	-22862
1st Coef(*100):	263434	131353
2nd Coef(*100):	5070	1477
H2O adj(mg/l*10k):	569	590

***** Printer Test End *****


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006677
Location = TOXL 8164.14.00 09/16
06/10/2022 09:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:39
02 Std. Sol.	0.017	09:40
03 Room Air	0.000	09:40
04 Std. Sol.	0.018	09:41
05 Room Air	0.000	09:42
06 Std. Sol.	0.018	09:42
07 Room Air	0.000	09:43

08 Sim Temp = 34.0°C

Simul Ser No = MP3060
Std Sol No = 202201A
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006677
Location = TOXL 8164.14.00 09/16
06/10/2022 09:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:49
02 Std. Sol.	0.297	09:50
03 Room Air	0.000	09:51
04 Std. Sol.	0.300	09:51
05 Room Air	0.000	09:52
06 Std. Sol.	0.301	09:53
07 Room Air	0.000	09:53

08 Sim Temp = 34.0°C

Simul Ser No = MP3064
Std Sol No = 202012A
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *LAP High Ac Check 0.300*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006677
Location = TOXL 8164.14.00 09/16
06/10/2022 09:54

Test	AC	Time
01 Room Air	0.000	09:55
02 *Subject Test	INT*	09:55
03 Room Air	0.000	09:56

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 20
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: ~~High AC Check~~
Interferent Check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006677
Location = TOXL 8164.14.00 09/16
06/10/2022 09:56

Test	AC	Time
01 Diagnostic	OK	09:57
02 Room Air	0.000	09:58
03 *Subject Test	RFI*	09:58
04 Room Air	0.000	09:58

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 20
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006677
Location = TOXL 8164.14.00 09/16
06/10/2022 09:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:59
02 Std. Gas	0.080	10:00
03 Room Air	0.000	10:00
04 Std. Gas	0.081	10:01
05 Room Air	0.000	10:01
06 Std. Gas	0.080	10:01
07 Room Air	0.000	10:02

Lot No = 02621080A1
Cyl No = 20
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006677
Location = TOXL 8164.14.00 09/16
06/10/2022 10:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:03
02 Std. Gas	0.080	10:03
03 Room Air	0.000	10:04
04 Std. Gas	0.080	10:04
05 Room Air	0.000	10:05
06 Std. Gas	0.080	10:05
07 Room Air	0.000	10:06

Lot No = 02621080A1
Cyl No = 20
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check # 4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006677
Location = TOXL 8164.14.00 09/16
06/10/2022 10:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:07
02 Std. Gas	0.081	10:07
03 Room Air	0.000	10:08
04 Std. Gas	0.081	10:08
05 Room Air	0.000	10:09
06 Std. Gas	0.080	10:09
07 Room Air	0.000	10:09

Lot No = 02621080A1

Cyl No = 20

Exp Date = 03/05/2023

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check # 7-9*