



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Nathan C Moeen</i>		
Location <i>GFCC</i>	Intoxilyzer® Serial Number <i>80-006507</i>	
Gas Lot Number <i>Z6021080A1</i>	Gas Cylinder Number <i>3</i>	Gas Expiration Date <i>10/05/2023</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081)..
 A. 0. 080 AC
 B. 0. 079 AC
 C. 0. 080 AC
- 3. Complete the-Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

*21 July 22
no
see note below*

Chemical Test Operator Signature <i>Nathan Moeen</i>	Date <i>6-10-22</i>
Reviewed By (Crime Laboratory Use Only) <i>Janelle Pertschiller</i>	Date <i>11 June 22</i>

Note: I called Nathan Moeen on 21 July 22 to verify a form 120-G was placed by the instrument at the gas cylinder installation. He has said yes. for 21 July 22

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = GFSO 8164.14.00 09/16
06/10/2022 20:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:56
02 Std. Gas	0.080	20:57
03 Room Air	0.000	20:57
04 Std. Gas	0.079	20:57
05 Room Air	0.000	20:58
06 Std. Gas	0.080	20:58
07 Room Air	0.000	20:59


Lot No = 26021080A1

Cyl No = 3

Exp Date = 10/05/2023

County = 18

Oper No. = 100452


Operator Signature
NATHAN MOEN

Remarks:

Form 106-I8000