

**CHAIN OF CUSTODY  
BREATH ALCOHOL INSTRUMENT**

Instrument: Intoxilyzer 8000 Serial Number: 80-006674

Agency being assigned instrument: NDOAG

Date 6/13/22

<u>Transferred from:</u>	<u>Transferred to:</u>
Name (print) <u>DION STEINLEY</u>	Name (print) <u>Deanna Dailey</u>
Signature <u>[Signature]</u>	Signature <u>[Signature]</u>
Agency <u>KILDEER PD</u>	Agency <u>Kildur PD</u>

Reason for transfer/remarks:  
Transferred to DAG - Crime Lab for annual inspection.

Note: High calibration check was low but may have been due to a bad simulator soln. pp 061422

Replaced to 80-004938  
ADL  
061322

USPS \_\_\_\_\_ UPS \_\_\_\_\_ FedEx \_\_\_\_\_ Other \_\_\_\_\_

Date Sending Instrument: \_\_\_\_\_ / \_\_\_\_\_ / \_\_\_\_\_ Air Bill No. \_\_\_\_\_