

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006689

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1000 psi Regulator: 1600 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3060 Lot #: 2022014 Exp. Date: 1/11/24
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12/8/22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1CS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 0262-108A1 Cyl #: 4 Exp. Date: 3/5/23

Test 1 <u>0.078</u>	Test 4 <u>0.078</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.078</u>	Test 8 <u>0.079</u>
Test 3 <u>0.078</u>	Test 6 <u>0.079</u>	Test 9 <u>0.078</u>
Average <u>0.078</u>		

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube o-ring. Replaced regulator due to low dry calibration gas initial readings. Purging the instrument improved the dry calibration gas readings. May have been caused by carryover from the interferent check.

Instrument is acceptable to be used in the field.

Jamilla Putschiller
Breath Analyst Signature

6/14/22

Date

[Signature]
Reviewed by

06/15/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
06/14/2022 08:56

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34921080A1
Standard Cyl #? 56
Standard Expiration? 02/05/2024
Oper No? 132128

Flow Cal. Date: 05/29/2020
Slope 643
Intercept -576193

IR Calibration Date: 05/29/2020

	3um	9um
0th Coef(*100):	-25870	-28277
1st Coef(*100):	287704	139899
2nd Coef(*100):	26	744
H2O adj(mg/l*10k):	293	328

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
06/14/2022 09:00

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:01
02 Std. Sol.	0.020	09:02
03 Room Air	0.000	09:02
04 Std. Sol.	0.020	09:03
05 Room Air	0.000	09:03
06 Std. Sol.	0.020	09:04
07 Room Air	0.000	09:05

08 Sim Temp = 34.0°C

Simul Ser No = MP3060
Std Sol No = 202201A
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low Ac Check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
06/14/2022 09:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:07
02 Std. Sol.	0.299	09:08
03 Room Air	0.000	09:09
04 Std. Sol.	0.300	09:10
05 Room Air	0.000	09:10
06 Std. Sol.	0.301	09:11
07 Room Air	0.000	09:12


08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 202012A

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC Check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
06/14/2022 09:16

Test	AC	Time
01 Diagnostic	OK	09:17
02 Room Air	0.000	09:17
03 *Subject Test	RFI*	09:18
04 Room Air	0.000	09:18

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 4
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
06/14/2022 21:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:09
02 Std. Gas	0.078	21:09
03 Room Air	0.000	21:09
04 Std. Gas	0.079	21:10
05 Room Air	0.000	21:10
06 Std. Gas	0.078	21:11
07 Room Air	0.000	21:11

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
06/14/2022 21:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:12
02 Std. Gas	0.078	21:12
03 Room Air	0.000	21:13
04 Std. Gas	0.078	21:13
05 Room Air	0.000	21:14
06 Std. Gas	0.079	21:14
07 Room Air	0.000	21:15

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
06/14/2022 21:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:16
02 Std. Gas	0.079	21:16
03 Room Air	0.000	21:17
04 Std. Gas	0.079	21:17
05 Room Air	0.000	21:17
06 Std. Gas	0.078	21:18
07 Room Air	0.000	21:18

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*