

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6688

Location: TOX L

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 994 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP 3060 Lot #: 202201A Exp. Date: 1/11/24
- b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: MP 90 Lot #: 202012A Exp. Date: 12/8/20
MP 3064

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR 5114 Lot #: ICSG Exp. Date: NA

Display reads "Interferent Detect"

3. RFI Check (CMS Mode) Display reads "RFI Detect"

4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC

Gas Cyl Lot #: 026210B0A1 Cyl #: 06 4 Exp. Date: 3/5/2023

Test 1	<u>0.079</u>	Test 4	<u>0.080</u>	Test 7	<u>0.080</u>
Test 2	<u>0.079</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.081</u>	Test 9	<u>0.080</u> ^{20 July 22} <u>0.081</u>
Average <u>0.080</u>					

C. Remarks/Maintenance: Annual inspection.

Replaced o-ring on breath tube nose.

Instrument is acceptable to be used in the field.

Janella Putschiller
Breath Analyst Signature

6/16/22
Date

[Signature]
Reviewed by

06/16/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
06/16/2022 09:42

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* () _ + ?

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* () _ + ?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 34
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 09/07/2016
Slope 698
Intercept -642832

IR Calibration Date: 09/07/2016
 3um 9um

0th Coef(*100): -10822 -23152
1st Coef(*100): 262468 134492
2nd Coef(*100): 3796 1436
H2O adj(mg/l*10k): 665 568

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
06/16/2022 09:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:52
02 Std. Sol.	0.018	09:53
03 Room Air	0.000	09:54
04 Std. Sol.	0.018	09:54
05 Room Air	0.000	09:55
06 Std. Sol.	0.019	09:56
07 Room Air	0.000	09:56

08 Sim Temp = 34.0°C

Simul Ser No = MP3060
Std Sol No = 202201A
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC Check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
06/16/2022 10:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:14
02 Std. Gas	0.080	10:14
03 Room Air	0.000	10:15
04 Std. Gas	0.080	10:15
05 Room Air	0.000	10:16
06 Std. Gas	0.081	10:16
07 Room Air	0.000	10:17

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry calibration check #4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
06/16/2022 10:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:18
02 Std. Gas	0.080	10:19
03 Room Air	0.000	10:19
04 Std. Gas	0.080	10:19
05 Room Air	0.000	10:20
06 Std. Gas	0.081	10:20
07 Room Air	0.000	10:21

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check 7-9*