



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-007086	Instrument Location Lisbon
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Install after Yearly Maintenance</u>	

- Check When Done:
- 1. Surge Protector Installed/Property Grounded.
 - 2. Telephone Line Connected to Intoxilyzer® 8000.
 - 3. Breath Tube Heated.
 - 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
 - 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
 - 6. Run Tests:
 - A. Print Test (Level1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
 - 7. Repair and/or Maintenance Performed (if any): None
 - 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
 - 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
 - 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date 6-20-22
---	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Janelle Pertschiller</i>	Certified Date 06/24/22
--	-----------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = LISB 8164.16.00 09/18
06/20/2022 16:41

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34921080A1
Standard Cyl #? 60
Standard Expiration? 02/05/2024
Oper No? 020607

Flow Cal. Date: 10/16/2018
Slope 688
Intercept -653049

IR Calibration Date: 10/16/2018
 3um 9um

0th Coef (*100): -47466 -23263
1st Coef (*100): 233222 130299
2nd Coef (*100): 3639 1618
H2O adj(mg/l*10k): 1280 608

***** Printer Test End *****



Operator Signature
TIMOTHY JORDAN

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = LISB 8164.16.00 09/18
06/20/2022 16:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:46
02 Std. Gas	0.080	16:46
03 Room Air	0.000	16:47
04 Std. Gas	0.081	16:47
05 Room Air	0.000	16:48
06 Std. Gas	0.081	16:48
07 Room Air	0.000	16:49

Lot No = 34921080A1
Cyl No = 60
Exp Date = 02/05/2024
County = 37 Oper No. = 020607



Operator Signature
TIMOTHY JORDAN

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = LISB 8164.16.00 09/18
06/20/2022 16:51

Test	AC	Time
01 Room Air	RFI*	16:52
02 Room Air	0.000	16:52

*Invalid Test
Inhibited - RFI

Sub Name = JORDAN, TIMOTHY ALLEN
Sub DOB = 01/11/1983
Sub Sex = Male Weight = 335
Test = OTH Cit = NA
Dr. Lic. = ND/JOR839841
Lot No = 34921080A1
Cyl No = 60
Expiration Date = 02/05/2024
County = 37 Oper No. = 020607



Operator Signature
TIMOTHY JORDAN

Remarks:

RFI Test

Form 106-I8000