

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006512 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 723 psi Regulator: 750 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP5290 Lot #: 21340 Exp. Date: 8/24/23
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3067 Lot #: 21420 Exp. Date: 10/20/23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1656 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 31 Exp. Date: 3/5/23

Test 1	<u>0.079</u>	Test 4	<u>0.079</u>	Test 7	<u>0.080</u>
Test 2	<u>0.079</u>	Test 5	<u>0.079</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average <u>0.080</u>					

C. Remarks/Maintenance: Annual inspection.
Replaced breath tube o-ring.

Instrument is acceptable to be used in the field.

Stinella Putschiller
Breath Analyst Signature

6/22/22
Date

[Signature]
Reviewed by

062322
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.14.00 09/16
06/22/2022 15:09

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05620080A1
Standard Cyl #? 3
Standard Expiration? 04/05/2022
Oper No? 132128

Flow Cal. Date: 08/20/2015
Slope 680
Intercept -660856

IR Calibration Date: 08/21/2015
 3um 9um

0th Coef(*100): -23051 -19261
1st Coef(*100): 266836 137135
2nd Coef(*100): 3974 1466
H2O adj(mg/l*10k): 785 487

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.14.00 09/16
06/22/2022 15:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:26
02 Std. Sol.	0.299	15:28
03 Room Air	0.000	15:28
04 Std. Sol.	0.301	15:29
05 Room Air	0.000	15:30
06 Std. Sol.	0.301	15:30
07 Room Air	0.000	15:31


08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 21420

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.14.00 09/16
06/22/2022 15:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:42
02 Std. Gas	0.079	15:43
03 Room Air	0.000	15:43
04 Std. Gas	0.079	15:44
05 Room Air	0.000	15:44
06 Std. Gas	0.080	15:45
07 Room Air	0.000	15:45

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.14.00 09/16
06/22/2022 15:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:53
02 Std. Gas	0.080	15:53
03 Room Air	0.000	15:54
04 Std. Gas	0.080	15:54
05 Room Air	0.000	15:55
06 Std. Gas	0.080	15:55
07 Room Air	0.000	15:56

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*