

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-0079 Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 696 psi Regulator: 700 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
    - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP5240 Lot #: 21340 Exp. Date: 8/24/23
    - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3067 Lot #: 21420 Exp. Date: 10/20/23
  - Configure dry gas standard for the following tests (Level 1,S).
  2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR5114 Lot #: 1CS6 Exp. Date: NA  
 Display reads "Interferent Detect"
  3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
  4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 02621080A1 Cyl #: 31 Exp. Date: 3/5/23

Test 1	<u>0.080</u>	Test 4	<u>0.079</u>	Test 7	<u>0.079</u>
Test 2	<u>0.079</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.079</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average	<u>0.079</u> $\approx$ <u>0.080</u>				

C. Remarks/Maintenance: Annual inspection.  
Replaced breath tube o-ring

Instrument is acceptable to be used in the field.

Janele Pertschiller  
Breath Analyst Signature

6/23/22  
Date

[Signature]  
Reviewed by

062822  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006679  
Location = TOXL                    8164.14.00 09/16  
06/23/2022                            09:49

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&\*()\_+?

Current Instrument Setup

Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer(Y/N)?                No  
Display Volume?                    No  
Disable On Memfull?                Yes  
# of Print Copies?                   1  
Select Std (D/W/I)?                Dry  
Standard Value?                    0.080  
Standard Lot #?                    26021080A1  
Standard Cyl #?                    18  
Standard Expiration?                10/05/2023  
Oper No?                            132128

Flow Cal. Date:                    09/06/2016  
Slope                                701  
Intercept                            -682926

IR Calibration Date:                09/06/2016  
                                          3um                                9um

-----  
0th Coef(\*100):            -18983                            -23165  
1st Coef(\*100):            280986                            134481  
2nd Coef(\*100):            2443                                1290  
H2O adj(mg/l\*10k):        569                                553

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006679  
Location = TOXL                    8164.14.00 09/16  
06/23/2022                            09:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:51
02 Std. Sol.	0.018	09:52
03 Room Air	0.000	09:53
04 Std. Sol.	0.017	09:53
05 Room Air	0.000	09:54
06 Std. Sol.	0.018	09:54
07 Room Air	0.000	09:55

08 Sim Temp = 34.0°C

Simul Ser No = MP5290  
Std Sol No = 21340  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low AC Check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006679  
Location = TOXL      8164.14.00 09/16  
06/23/2022      10:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:07
02 Std. Sol.	0.291	10:08
03 Room Air	0.000	10:08
04 Std. Sol.	0.292	10:09
05 Room Air	0.000	10:10
06 Std. Sol.	0.293	10:10
07 Room Air	0.000	10:11

08 Sim Temp = 34.0°C

Simul Ser No = MP3067  
Std Sol No = 21420  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High AC Check 0.300*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006679  
Location = TOXL                    8164.14.00 09/16  
06/23/2022                            10:14

Test	AC	Time
01 Room Air	0.000	10:15
02 *Subject Test	INT*	10:15
03 Room Air	0.000	10:16

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                            Weight = NA  
Test = DUI                            Cit = INTERFERENT CHK  
Dr. Lic. = ND/TES989643  
Lot No = 02621080A1  
Cyl No = 31  
Expiration Date = 03/05/2023  
County = 08                            Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006679  
Location = TOXL                    8164.14.00 09/16  
06/23/2022                            10:16

Test	AC	Time
01 Diagnostic	OK	10:17
02 Room Air	0.000	10:17
03 *Subject Test	RFI*	10:18
04 Room Air	0.000	10:18

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                      Weight = NA  
Test = DUI                            Cit = RFI CHK  
Dr. Lic. = ND/TES989643  
Lot No = 02621080A1  
Cyl No = 31  
Expiration Date = 03/05/2023  
County = 08                            Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI Check*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006679  
Location = TOXL      8164.14.00 09/16  
06/23/2022      10:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:29
02 Std. Gas	0.080	10:29
03 Room Air	0.000	10:30
04 Std. Gas	0.079	10:30
05 Room Air	0.000	10:31
06 Std. Gas	0.079	10:31
07 Room Air	0.000	10:32

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *dry calibration check # 1-3*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006679  
Location = TOXL                    8164.14.00 09/16  
06/23/2022                            10:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:33
02 Std. Gas	0.079	10:33
03 Room Air	0.000	10:34
04 Std. Gas	0.080	10:34
05 Room Air	0.000	10:35
06 Std. Gas	0.080	10:35
07 Room Air	0.000	10:36

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006679  
Location = TOXL                    8164.14.00 09/16  
06/23/2022                            10:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:39
02 Std. Gas	0.079	10:40
03 Room Air	0.000	10:40
04 Std. Gas	0.080	10:41
05 Room Air	0.000	10:41
06 Std. Gas	0.080	10:41
07 Room Air	0.000	10:42

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry calibration check #7-9*