

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006509 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 951 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP3070 Lot #: 22CP41184 Exp. Date: 3/1/24
 - b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.400 AC
Sim. Ser #: MP3065 Lot #: 202103E Exp. Date: 3/24/23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: 1656 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02671080A1 Cyl #: 16 Exp. Date: 3/5/23

Test 1 <u>0.079</u>	Test 4 <u>0.078</u>	Test 7 <u>0.079</u>
Test 2 <u>0.078</u>	Test 5 <u>0.078</u>	Test 8 <u>0.078</u>
Test 3 <u>0.079</u>	Test 6 <u>0.078</u>	Test 9 <u>0.079</u>
Average <u>0.078</u>		

C. Remarks/Maintenance: Annual inspection.
Breath tube o-ring was replaced. Replaced external battery

Instrument is acceptable to be used in the field.

Janelle Putschiller
Breath Analyst Signature

6/23/22
Date

[Signature]
Reviewed by

062422
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
06/23/2022 11:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:38
02 Std. Sol.	0.020	11:40
03 Room Air	0.000	11:40
04 Std. Sol.	0.020	11:41
05 Room Air	0.000	11:42
06 Std. Sol.	0.020	11:42
07 Room Air	0.000	11:43

08 Sim Temp = 34.0°C

Simul Ser No = MP3070
Std Sol No = 22CP41184
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC Check 0.020*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
06/23/2022 11:51

Test	AC	Time
01 Room Air	0.000	11:51
02 *Subject Test	INT*	11:52
03 Room Air	0.000	11:53

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CHK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 16
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
06/23/2022 11:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:58
02 Std. Gas	0.079	11:58
03 Room Air	0.000	11:59
04 Std. Gas	0.078	11:59
05 Room Air	0.000	12:00
06 Std. Gas	0.079	12:00
07 Room Air	0.000	12:01

Lot No = 02621080A1
Cyl No = 16
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
06/23/2022 12:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:02
02 Std. Gas	0.078	12:02
03 Room Air	0.000	12:02
04 Std. Gas	0.078	12:03
05 Room Air	0.000	12:03
06 Std. Gas	0.078	12:04
07 Room Air	0.000	12:04

Lot No = 02621080A1
Cyl No = 16
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

