



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-006509	Instrument Location Grand Forks AFB
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 24 Jun 22
---	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Jeanette Pertscheller	Certified Date 09/08/2022
---	-------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = GFAF 8164.14.00 09/16
06/24/2022 14:38

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 12
Standard Expiration? 03/05/2023
Oper No? 132924

Flow Cal. Date: 06/11/2019
Slope 445
Intercept 431160

IR Calibration Date: 06/11/2019

	3um	9um
0th Coef(*100):	-27266	-14145
1st Coef(*100):	276865	139368
2nd Coef(*100):	2326	1158
H2O adj(mg/l*10k):	584	288

***** Printer Test End *****



Operator Signature
ZACHARY BRODBECK

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = GFAF 8164.14.00 09/16
06/24/2022 14:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:39
02 Std. Gas	0.078	14:40
03 Room Air	0.000	14:40
04 Std. Gas	0.079	14:41
05 Room Air	0.000	14:41
06 Std. Gas	0.079	14:41
07 Room Air	0.000	14:42

Lot No = 02621080A1
Cyl No = 12
Exp Date = 03/05/2023
County = 65 Oper No. = 132924


Operator Signature
ZACHARY BRODBECK

Remarks:

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = GFAP 8164.14.00 09/16
06/24/2022 14:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI**	14:47

*Sequence Aborted
**Invalid Test
Inhibited - RFI

Lot No = 02621080A1
Cyl No = 12
Exp Date = 03/05/2023
County = NA Oper No. = NA


Operator Signature
Zachary Brodbeck 09/08/22 JRP

Remarks:

Form 106-I8000

Note: operator information was not added. I contacted Grand Forks AFB several times to update this information. They did not return my emails. 09/08/22 JRP