



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59281 (06/2018)

Serial Number 80-006512	Instrument Location Carrington PD (CARR)
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 6-25-22
-------------------------------	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date 27 June 22
---------------------------	-------------------------------------

I certify that this document is a true and correct duplicate of the original.
 Name Nickolas Oster 4281
 Date 6-25-22
 Police Department
 Carrington, N.D.

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
06/25/2022 21:03

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 35
Standard Expiration? 10/05/2023
Oper No? 131838

Flow Cal. Date: 08/20/2015
Slope 680
Intercept -660856

IR Calibration Date: 08/21/2015
3um 9um

	3um	9um
0th Coef(*100):	-23051	-19261
1st Coef(*100):	266836	137135
2nd Coef(*100):	3974	1466
H2O adj(mg/l*10k):	785	487

***** Printer Test End *****



Operator Signature
NICKOLAS OSTER

Remarks:

Print test

I certify that this document
is a true and correct duplicate
of the original.
Name Nickolas Oster 4281
Date 6-25-22
Police Department
Carrington, N.D.

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
06/25/2022 21:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:04
02 Std. Gas	0.081	21:05
03 Room Air	0.000	21:05
04 Std. Gas	0.080	21:06
05 Room Air	0.000	21:06
06 Std. Gas	0.080	21:06
07 Room Air	0.000	21:07

Lot No = 26021080A1
Cyl No = 35
Exp Date = 10/05/2023
County = 16 Oper No. = 131838



Operator Signature
NICKOLAS OSTER

Remarks:

ACA Test

Form 106-I8000

I certify that this document
is a true and correct duplicate
of the original

Name Nickolas Oster 4281
Date 6-25-22

Police Department
Carrington, N.D.

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
06/25/2022 21:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	21:11
02 Room Air	0.000	21:11

*Invalid Test
Inhibited - RFI

Lot No = 26021080A1
Cyl No = 35
Exp Date = 10/05/2023
County = 16

Oper No. = 131838



Operator Signature
NICKOLAS OSTER

Remarks:

RFI Test

Form 106-I8000

I certify that this document
is a true and correct duplicate
of the original.

Name Nickolas Oster 4081
Date 6-25-22

Police Department
Carrington, N.D.