

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006691

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 881 psi Regulator: 850 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP5290 Lot #: 21340 Exp. Date: 8/24/23
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3067 Lot #: 2140911 Exp. Date: 10/20/23
21420

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1C56 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 031 Exp. Date: 3/5/2023

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: Annual inspection.

Breath tube o-ring replaced.

Instrument is acceptable to be used in the field.

Janelle Pertschiller
Breath Analyst Signature

06/27/2022
Date

[Signature]
Reviewed by

062922
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = TOXL 8164.14.00 09/16
06/27/2022 12:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:22
02 Std. Sol.	0.018	12:23
03 Room Air	0.000	12:23
04 Std. Sol.	0.019	12:24
05 Room Air	0.000	12:24
06 Std. Sol.	0.019	12:25
07 Room Air	0.000	12:26


08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks:

Low AC check 0.020

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = TOXL 8164.14.00 09/16
06/27/2022 12:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:36
02 Std. Sol.	0.297	12:37
03 Room Air	0.000	12:38
04 Std. Sol.	0.298	12:38
05 Room Air	0.000	12:39
06 Std. Sol.	0.298	12:40
07 Room Air	0.000	12:40

08 Sim Temp = 34.0°C

Simul Ser No = MP3067
Std Sol No = 21420
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = TOXL 8164.14.00 09/16
06/27/2022 12:43

Test	AC	Time
01 Room Air	0.000	12:43
02 *Subject Test	INT*	12:44
03 Room Air	0.000	12:45

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 31
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = TOXL 8164.14.00 09/16
06/27/2022 12:45

Test	AC	Time
01 Diagnostic	OK	12:46
02 Room Air	0.000	12:47
03 *Subject Test	RFI*	12:48
04 Room Air	0.000	12:48

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 31
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = TOXL 8164.14.00 09/16
06/27/2022 12:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:50
02 Std. Gas	0.081	12:50
03 Room Air	0.000	12:51
04 Std. Gas	0.081	12:51
05 Room Air	0.000	12:51
06 Std. Gas	0.081	12:52
07 Room Air	0.000	12:52

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = TOXL 8164.14.00 09/16
06/27/2022 12:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:53
02 Std. Gas	0.081	12:53
03 Room Air	0.000	12:54
04 Std. Gas	0.081	12:54
05 Room Air	0.000	12:55
06 Std. Gas	0.081	12:55
07 Room Air	0.000	12:56

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128

Janelle Portscheller
Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = TOXL 8164.14.00 09/16
06/27/2022 12:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:57
02 Std. Gas	0.081	12:58
03 Room Air	0.000	12:58
04 Std. Gas	0.081	12:58
05 Room Air	0.000	12:59
06 Std. Gas	0.081	12:59
07 Room Air	0.000	13:00

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*