

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006687

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 918 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
    - a.  Low AC (Level 1,C): Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP3070 Lot #: 22CP41184 Exp. Date: 3/1/24
    - b.  High AC (Level 1,C): Known Value  $\geq$  0.25 AC: 0.400 AC  
Sim. Ser #: MP3065 Lot #: 202103E Exp. Date: 3/24/23

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3847 Lot #: 1CS6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.08 AC  
Gas Cyl Lot #: 34921080A1 Cyl #: 44 Exp. Date: 2/5/24

Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: Annual inspection

Replaced breath tube o-ring.

Instrument is acceptable to be used in the field.

Janelle Pertschiller  
Breath Analyst Signature

6/29/2022

Date

Reviewed by [Signature]

06/30/22

Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006687  
Location = TOXL      8164.14.00 09/16  
06/29/2022      10:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:10
02 Std. Sol.	0.018	10:12
03 Room Air	0.000	10:13
04 Std. Sol.	0.018	10:13
05 Room Air	0.000	10:14
06 Std. Sol.	0.018	10:14
07 Room Air	0.000	10:15

08 Sim Temp = 34.0°C

Simul Ser No = MP3070  
Std Sol No = 22CP41184  
County = 08      Oper No. = 132128

*Janelle Portscheller*

Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low Ac check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006687  
Location = TOXL      8164.14.00 09/16  
06/29/2022      10:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:26
02 Std. Sol.	0.394	10:29
03 Room Air	0.000	10:30
04 Std. Sol.	0.395	10:30
05 Room Air	0.000	10:31
06 Std. Sol.	0.396	10:32
07 Room Air	0.000	10:32


08 Sim Temp = 34.0°C

Simul Ser No = MP3065

Std Sol No = 202103E

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High AC check 6.400*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006687  
Location = TOXL                    8164.14.00 09/16  
06/29/2022                            10:39

Test	AC	Time
01 Diagnostic	OK	10:40
02 Room Air	0.000	10:40
03 *Subject Test	RFI*	10:41
04 Room Air	0.000	10:41

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHK  
Dr. Lic. = ND/DIS821456  
Lot No = 34921080A1  
Cyl No = 44  
Expiration Date = 02/05/2024  
County = 08                            Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI check*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006687  
Location = TOXL      8164.14.00 09/16  
06/29/2022      10:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:42
02 Std. Gas	0.080	10:43
03 Room Air	0.000	10:43
04 Std. Gas	0.081	10:44
05 Room Air	0.000	10:44
06 Std. Gas	0.080	10:44
07 Room Air	0.000	10:45


Lot No = 34921080A1

Cyl No = 44

Exp Date = 02/05/2024

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #13*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006687  
Location = TOXL      8164.14.00 09/16  
06/29/2022      10:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:46
02 Std. Gas	0.081	10:47
03 Room Air	0.000	10:47
04 Std. Gas	0.081	10:48
05 Room Air	0.000	10:48
06 Std. Gas	0.081	10:48
07 Room Air	0.000	10:49

Lot No = 34921080A1  
Cyl No = 44  
Exp Date = 02/05/2024  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check #4-6*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006687  
Location = TOXL      8164.14.00 09/16  
06/29/2022      10:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:51
02 Std. Gas	0.081	10:52
03 Room Air	0.000	10:52
04 Std. Gas	0.081	10:52
05 Room Air	0.000	10:53
06 Std. Gas	0.081	10:53
07 Room Air	0.000	10:54

Lot No = 34921080A1  
Cyl No = 44  
Exp Date = 02/05/2024  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check #7-9*