

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006513

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 999 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP 3060 Lot #: 202201A Exp. Date: V11/24
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP 3064 Lot #: 2020 6A Exp. Date: 12/8/22

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7345 Lot #: 1CS 7 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 3492.1080A1 Cyl #: 049 Exp. Date: 2/5/24

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.082</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube O-ring.  
Left instrument in maintain time for annual inspection.  
Replaced external battery.

Instrument is acceptable to be used in the field.

Janelle Pertschiller  
Breath Analyst Signature

06/30/2022  
Date

[Signature]  
Reviewed by

070122  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006513  
Location = TOXL                    8164.13.00 06/09  
06/30/2022                            11:37

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer(Y/N)?                No  
Display Volume?                    No  
Disable On Memfull?                Yes  
# of Print Copies?                  1  
Select Std (D/W/I)?                Dry  
Standard Value?                    0.080  
Standard Lot #?                    26021080A1  
Standard Cyl #?                    14  
Standard Expiration?                10/05/2023  
Oper No?                            132128

Flow Cal. Date:                    08/20/2015  
Slope                                680  
Intercept                            -696037

IR Calibration Date:                08/21/2015  
    3um                                9um

-----  
0th Coef(\*100):                    -14436                                -13610  
1st Coef(\*100):                    265490                                136147  
2nd Coef(\*100):                    4422                                    1710  
H2O adj(mg/l\*10k):                672                                    448

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006513  
Location = TOXL                    8164.13.00 06/09  
06/30/2022                            11:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:39
02 Std. Sol.	0.020	11:40
03 Room Air	0.000	11:41
04 Std. Sol.	0.021	11:41
05 Room Air	0.000	11:42
06 Std. Sol.	0.021	11:43
07 Room Air	0.000	11:43

08 Sim Temp = 34.0°C

Simul Ser No = MP3060

Std Sol No = 202201A

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006513  
Location = TOXL      8164.13.00 06/09  
06/30/2022      11:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:45
02 Std. Sol.	0.301	11:46
03 Room Air	0.000	11:46
04 Std. Sol.	0.304	11:47
05 Room Air	0.000	11:47
06 Std. Sol.	0.303	11:48
07 Room Air	0.000	11:49

08 Sim Temp = 34.0°C

Simul Ser No = MP3064  
Std Sol No = 202012A  
County = 08      Oper No. = 132128

  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: High ~~at~~ AC check 0.300


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006513  
Location = TOXL                    8164.13.00 06/09  
06/30/2022                            11:50

Test	AC	Time
01 Room Air	0.000	11:50
02 *Subject Test	INT*	11:51
03 Room Air	0.000	11:52

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT CK  
Dr. Lic. = ND/DIS821456  
Lot No = 34921080A1  
Cyl No = 49  
Expiration Date = 02/05/2024  
County = 08                            Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-006513  
Location = TOXL                              8164.13.00 06/09  
06/30/2022    11:52

Test	AC	Time
01 Diagnostic	OK	11:53
02 Room Air	0.000	11:53
03 *Subject Test	RFI*	11:54
04 Room Air	0.000	11:54

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                              Weight = 150  
Test = DUI                                      Cit = RFI CHK  
Dr. Lic. = ND/DIS821456  
Lot No = 34921080A1  
Cyl No = 49  
Expiration Date = 02/05/2024  
County = 08                                      Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006513  
Location = TOXL                    8164.13.00 06/09  
06/30/2022                            11:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:55
02 Std. Gas	0.081	11:56
03 Room Air	0.000	11:56
04 Std. Gas	0.081	11:57
05 Room Air	0.000	11:57
06 Std. Gas	0.081	11:57
07 Room Air	0.000	11:58

Lot No = 34921080A1  
Cyl No = 49  
Exp Date = 02/05/2024  
County = 08                            Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006513  
Location = TOXL                    8164.13.00 06/09  
06/30/2022                            11:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:59
02 Std. Gas	0.081	11:59
03 Room Air	0.000	12:00
04 Std. Gas	0.081	12:00
05 Room Air	0.000	12:00
06 Std. Gas	0.081	12:01
07 Room Air	0.000	12:01

Lot No = 34921080A1  
Cyl No = 49  
Exp Date = 02/05/2024  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006513  
Location = TOXL                    8164.13.00 06/09  
06/30/2022                            12:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:03
02 Std. Gas	0.081	12:03
03 Room Air	0.000	12:04
04 Std. Gas	0.081	12:04
05 Room Air	0.000	12:04
06 Std. Gas	0.082	12:05
07 Room Air	0.000	12:05

Lot No = 34921080A1  
Cyl No = 49  
Exp Date = 02/05/2024  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration checks #7-9*