



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number 80-006507	Instrument Location Grand Forks County Corrections (GFSO)
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): Data file upload at State Lab
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 10/19/2022
--	--------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date 10/20/22
--	----------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006507  
Location = GFSO      8164.14.00 09/16  
10/19/2022      09:03

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? Yes  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 26021080A1  
Standard Cyl #? 3  
Standard Expiration? 10/05/2023  
Oper No? 020836

Flow Cal. Date: 08/20/2015  
Slope 659  
Intercept -658315

IR Calibration Date: 08/20/2015  
3um      9um

-----  
0th Coef(\*100):      -9452      -11722  
1st Coef(\*100):      258808      139588  
2nd Coef(\*100):      4476      1674  
H2O adj(mg/l\*10k): 499      376

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
DAVID STROMBERG

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006507  
Location = GFSO      8164.14.00 09/16  
10/19/2022      09:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:05
02 Std. Gas	0.082	09:05
03 Room Air	0.000	09:06
04 Std. Gas	0.082	09:06
05 Room Air	0.000	09:07
06 Std. Gas	0.082	09:07
07 Room Air	0.000	09:08

Lot No = 26021080A1  
Cyl No = 3  
Exp Date = 10/05/2023  
County = 18      Oper No. = 020836

  
Operator Signature  
DAVID STROMBERG

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006507  
Location = GFSO      8164.14.00 09/16  
10/19/2022      09:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	09:03
02 Room Air	0.000	09:04

\*Invalid Test  
Inhibited - RFI

Lot No = 26021080A1  
Cyl No = 3  
Exp Date = 10/05/2023  
County = 18      Oper No. = 020836

  
Operator Signature  
DAVID STROMBERG

Remarks: RFI Test