

**INTOXILYZER® 8000 INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-006673

Location: TUCL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 775 psi Regulator: 775 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP3070 Lot #: 22CP41184 Exp. Date: 3/1/24
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.400 AC  
Sim. Ser #: MP3065 Lot #: 202103E Exp. Date: 3/24/23

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR5114 Lot #: IC56 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 02621080A1 Cyl #: 2 Exp. Date: 3/5/23

Test 1	<u>0.082</u>	Test 4	<u>0.081</u>	Test 7	<u>0.082</u>
Test 2	<u>0.082</u>	Test 5	<u>0.082</u>	Test 8	<u>0.082</u>
Test 3	<u>0.081</u>	Test 6	<u>0.082</u>	Test 9	<u>0.082</u>
Average	<u>0.082</u>				

C. Remarks/Maintenance: Annual Inspection

Replaced battery. Replaced o-ring on breath hose. Replaced o-ring on simulator return.

Instrument is acceptable to be used in the field.

Janelle Pertscheller  
Breath Analyst Signature

4/6/22  
Date

[Signature]  
Reviewed by

04/22/22  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006673  
Location = TOXL                8164.14.00 09/16  
04/06/2022                      11:26

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            05620080A1  
Standard Cyl #?            33  
Standard Expiration?       04/05/2022  
Oper No?                    132128

Flow Cal. Date:            03/30/2017  
Slope                      697  
Intercept                  -667725

IR Calibration Date:       03/30/2017  
                                  3um                    9um

-----  
0th Coef(\*100):        -9136                -17613  
1st Coef(\*100):        267126              133333  
2nd Coef(\*100):        5438                1698  
H2O adj(mg/l\*10k):    684                 589

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006673  
Location = TOXL                    8164.14.00 09/16  
04/06/2022                            11:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:28
02 Std. Sol.	0.018	11:29
03 Room Air	0.000	11:29
04 Std. Sol.	0.018	11:30
05 Room Air	0.000	11:30
06 Std. Sol.	0.019	11:31
07 Room Air	0.000	11:32

08 Sim Temp = 34.0°C

Simul Ser No = MP3070  
Std Sol No = 22CP41184  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low Ac 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006673  
Location = TOXL      8164.14.00 09/16  
04/06/2022      11:33

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:34
02 Std. Sol.	0.396	11:35
03 Room Air	0.000	11:35
04 Std. Sol.	0.401	11:36
05 Room Air	0.000	11:37
06 Std. Sol.	0.403	11:37
07 Room Air	0.000	11:38


08 Sim Temp = 34.0°C

Simul Ser No = MP3065

Std Sol No = 202103E

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: High AC ~~0.000~~ 0.400

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006673  
Location = TOXL            8164.14.00 09/16  
04/06/2022            11:41

Test	AC	Time
01 Room Air	0.000	11:42
02 *Subject Test	INT*	11:43
03 Room Air	0.000	11:43

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female            Weight = 150  
Test = DUI            Cit = INTERFERENT CK  
Dr. Lic. = ND/DIS821456  
Lot No = 02621080A1  
Cyl No = 2  
Expiration Date = 03/05/2023  
County = 08            Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006673  
Location = TOXL                    8164.14.00 09/16  
04/06/2022                            11:43

Test	AC	Time
01 Diagnostic	OK	11:44
02 Room Air	0.000	11:45
03 *Subject Test	RFI*	11:45
04 Room Air	0.000	11:46

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CK  
Dr. Lic. = ND/DIS821456  
Lot No = 02621080A1  
Cyl No = 2  
Expiration Date = 03/05/2023  
County = 08                            Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI Check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006673  
Location = TOXL      8164.14.00 09/16  
04/06/2022      11:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:58
02 Std. Gas	0.082	11:58
03 Room Air	0.000	11:59
04 Std. Gas	0.082	11:59
05 Room Air	0.000	12:00
06 Std. Gas	0.081	12:00
07 Room Air	0.000	12:00

Lot No = 02621080A1  
Cyl No = 2  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

Janelle Portscheller  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006673  
Location = TOXL      8164.14.00 09/16  
04/06/2022      12:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:02
02 Std. Gas	0.081	12:02
03 Room Air	0.000	12:02
04 Std. Gas	0.082	12:03
05 Room Air	0.000	12:03
06 Std. Gas	0.082	12:04
07 Room Air	0.000	12:04

Lot No = 02621080A1  
Cyl No = 2  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check #4-6*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006673  
Location = TOXL      8164.14.00 09/16  
04/06/2022      12:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:08
02 Std. Gas	0.082	12:08
03 Room Air	0.000	12:09
04 Std. Gas	0.082	12:09
05 Room Air	0.000	12:10
06 Std. Gas	0.082	12:10
07 Room Air	0.000	12:10

Lot No = 02621080A1  
Cyl No = 2  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*