



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Wyatt E. Schilling</i>		
Location <i>SACO</i> <i>Sargent County Sheriff Office</i>	Intoxilyzer® Serial Number <i>80-006681</i>	
Gas Lot Number <i>34921080A1</i>	Gas Cylinder Number <i>057</i>	Gas Expiration Date <i>2/5/2024</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.079* AC
 - B. *0.079* AC
 - C. *0.078* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature 	Date <i>4/6/2022</i>
Reviewed By (Crime Laboratory Use Only) <i>Janelle Pertscheller</i>	Date <i>20 Apr 22</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = SACO 8164.14.00 09/16
04/06/2022 15:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:35
02 Std. Gas	0.079	15:36
03 Room Air	0.000	15:36
04 Std. Gas	0.079	15:37
05 Room Air	0.000	15:37
06 Std. Gas	0.078	15:37
07 Room Air	0.000	15:38

Lot No = 34921080A1
Cyl No = 57
Exp Date = 02/05/2024
County = 41 Oper No. = 132834


Operator Signature
WYATT SCHILLING

Remarks:

Form 106-I8000

***I certify that the foregoing is a true
and correct copy of the original
document on file in this office.***


SCSO