



OFFICE OF ATTORNEY GENERAL
Crime Laboratory Division
2641 East Main Avenue
P.O. Box 937
Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Benny Vonsett

Location: Mohl

Intoxilyzer Serial No.: 80-004198

Lot Number: 659358

Cylinder Number: 27

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 080 %

Test 2: 0. 081 %

Test 3: 0. 080 %

(ATTACH TEST RECORD)

Benny Vonsett 1-1-10
Chemical Test Operator Signature

31 Mar 2010
Date Cylinder Changed *Can not remember
if cylinder was
changed or not*

Deb Kashur
Reviewed By

31 Mar 2010
Date

Form 105-G
Toxicology Laboratory
Crime Laboratory Division

[Type text]

80-004198
77/59 FM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = ~~FOXL~~ *MotHL* DK 8164.13.00 06/09
01/01/2010 22:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:29
02 Std. Gas	0.080	22:29
03 Room Air	0.000	22:30
04 Std. Gas	0.081	22:30
05 Room Air	0.000	22:31
06 Std. Gas	0.080	22:31
07 Room Air	0.000	22:32

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 039909


Operator Signature
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-00 4198
7/8/79 *KL*

Intoxilyzer Test Record and Checklist
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
 North Dakota Model 8000 SN 80-004198
 Location = TFXL/Model B ✓ 8164.13.00 06/09
 01/01/2010 22:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:29
02 Std. Gas	0.080	22:29
03 Room Air	0.000	22:30
04 Std. Gas	0.081	22:30
05 Room Air	0.000	22:31
06 Std. Gas	0.080	22:31
07 Room Air	0.000	22:32


Lot No = 659358

Cyl No = 27

Exp Date = 09/03/2011

County = 38

Oper No. = 039909



Operator Signature
 BARRY A. VANNATTA

Remarks:

Form 106-18000

80-004198
 7/9/99 FWL