



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Lt. Tracy L. Meidinger #258

Location: UND-PD

Intoxilyzer Serial No.: 80-004206

Lot Number: 659358

Cylinder Number: 31

Expiration Date: 09-03-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.084 %

Test 2: 0.085 %

Test 3: 0.084 %

(ATTACH TEST RECORD)

Lt. Tracy L. Meidinger #258
 Chemical Test Operator Signature

01-02-2010
 Date Cylinder Changed

Deb Kashur (ACA check)
 Reviewed By

07 Jan 2010
 Date

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004206
 P. 1/2/2010

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = UNDP 8164.13.00 06/09
01/02/2010 20:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:41
02 Std. Gas	0.084	20:41
03 Room Air	0.000	20:42
04 Std. Gas	0.085	20:42
05 Room Air	0.000	20:42
06 Std. Gas	0.084	20:43
07 Room Air	0.000	20:43

Lot No = 659358
Cyl No = 31
Exp Date = 09/03/2011
County = 18 Oper No. = 119409

Tracy L. Meidinger #258
Operator Signature
TRACY L. MEIDINGER

Remarks:

ACA
01-02-2010

Form 106-I8000

80-004206
P. 2 of 2 YMA