



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Barry Vannatta</u>		
Location: <u>MOBILE</u>		
Intoxilyzer® Serial Number: <u>80-603069</u>		
Gas Lot Number: <u>27</u>	Gas Cylinder Number: <u>659358</u>	Expiration Date: <u>09-03-11</u>
<input type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.080</u> %		
Test 2: <u>0.080</u> %		
Test 3: <u>0.081</u> %		
(ATTACH TEST RECORD)		

Barry Vannatta
 Chemical Test Operator Signature

01-03-2011
 Date

Deb Kashur
 Reviewed By

10 Jan 2011
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = MOHL 8164.13.00 06/09
01/03/2011 10:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:57
02 Std. Gas	0.080	10:58
03 Room Air	0.000	10:58
04 Std. Gas	0.080	10:59
05 Room Air	0.000	10:59
06 Std. Gas	0.081	10:59
07 Room Air	0.000	11:00


Lot No = 659358

Cyl No = 27

Exp Date = 09/03/2011

County = 38

Oper No. = 039909


Operator Signature
BARRY VANNATTA

Remarks:

Form 106-I8000

80-003069
P. 82 of 82
JM