



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David M Wallace</u>		
Location: <u>Dickinson JEC</u>		
Intoxilyzer® Serial Number: <u>80-004947</u>		
Gas Lot Number: <u>2541008081</u>	Gas Cylinder Number: <u>20</u>	Expiration Date: <u>7-1-12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.079</u> %		
Test 2: <u>0.079</u> %		
Test 3: <u>0.079</u> %		
(ATTACH TEST RECORD)		

DM Wallace
Chemical Test Operator Signature

1-3-11
Date

Deb Kashur
Reviewed By

10 Jan 2011
Date

80-004947
P. 57 of 58
JEM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = DKPD 8164.13.00 06/09
01/03/2011 09:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:58
02 Std. Gas	0.079	09:58
03 Room Air	0.000	09:59
04 Std. Gas	0.079	09:59
05 Room Air	0.000	10:00
06 Std. Gas	0.079	10:00
07 Room Air	0.000	10:01

Lot No = 15410080A1
Cyl No = 20
Exp Date = 07/01/2012
County = 45 Oper No. = 068332


Operator Signature
DAVID M. WALLACE

Remarks:

Form 106-18000

80-004947
P. 58 of 58
FM