



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Mark Countess</i>		
Location <i>HARV</i>	Intoxilyzer® Serial Number <i>80-004950</i>	
Gas Lot Number <i>23411080A1</i>	Gas Cylinder Number <i>22</i>	Gas Expiration Date <i>10/1/13</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.080* AC
 - B. *0.080* AC
 - C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Mark Countess</i>	Date <i>1-4-12</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashum</i>	Date <i>03 Apr 11</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = HARV 8164.13.00 06/09
01/04/2012 23:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:31
02 Std. Gas	0.080	23:31
03 Room Air	0.000	23:32
04 Std. Gas	0.080	23:32
05 Room Air	0.000	23:32
06 Std. Gas	0.080	23:33
07 Room Air	0.000	23:33

Lot No = 23411080A1
Cyl No = 22
Exp Date = 10/01/2013
County = 52 Oper No. = 030819



Operator Signature
MARK A. COUNTESS

Remarks: