



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David J. Zeis</u>		
Location: <u>CRAVHER County SO Langdon</u>		
Intoxilyzer® Serial Number: <u>80-004937</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>22</u>	Expiration Date: <u>09/03/2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>078</u> %		
Test 2: 0. <u>080</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

David J. Zeis
 Chemical Test Operator Signature

01/05/2011
 Date

Deb Kashner
 Reviewed By

10 Jan 2011
 Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = LANG 8164.13.00 06/09
01/05/2011 10:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:06
02 Std. Gas	0.078	10:06
03 Room Air	0.000	10:07
04 Std. Gas	0.080	10:07
05 Room Air	0.000	10:08
06 Std. Gas	0.079	10:08
07 Room Air	0.000	10:09

Lot No = 659358
Cyl No = 22
Exp Date = 09/03/2011
County = 10 Oper No. = 067518



Operator Signature
DAVID J. ZEIS

Remarks:

Form 106-I8000

80-004937
P. 55 of 55
YFH