

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David M Wallace

Location: Dickinson JCC

Intoxilyzer Serial No.: 80-003066

Lot Number: 65-9358

Cylinder Number: 23

Expiration Date: 9-3-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 082 %

Test 2: 0. 083 %

Test 3: 0. 083 %

(ATTACH TEST RECORD)

David M Wallace
Chemical Test Operator Signature

1-15-2010 ^{DK} 01/06/2010
Date

Deb Kashur
Reviewed By

22 Jan 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = DKPD 8164.13.00 06/09
01/06/2010 14:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:54
02 Std. Gas	0.082	14:54
03 Room Air	0.000	14:55
04 Std. Gas	0.083	14:55
05 Room Air	0.000	14:56
06 Std. Gas	0.083	14:56
07 Room Air	0.000	14:57

Lot No = 659358
Cyl No = 23
Exp Date = 09/03/2011
County = 45 Oper No. = 068332



Operator Signature
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-003066
P. 2 of 2 FM