

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-004189

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5144 Lot No. Rep AC 0.019 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5187 Lot No. Rep AC 0.285 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR5114 Lot No. Rep AC 0.100 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 659358 Cyl No. 34 Exp. Date 03 Sept 2011

Test 1 0.079 Test 4 0.079 Test 7 0.079

Test 2 0.079 Test 5 0.079 Test 8 0.079

Test 3 0.079 Test 6 0.079 Test 9 0.079

Average 0.079

L. Remarks/Maintenance record: Software upgrade 8/6/13 Annual inspection

16 Jan 2010
Date

Deb Kashur
Field Inspector's Signature

14 January 2010
Date

Maryn Pleasom
Reviewed by

80-004189
6/3/11 YRM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.13.00 06/09
01/06/2010 08:54

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659358
Standard Cyl #? 34
Standard Expiration? 09/03/2011
Oper No? 777777

Flow Cal. Date: 06/03/2009
Slope 666
Intercept -550210

IR Calibration Date: 01/05/2010

	3um	9um
0th Coef(*100):	-13528	-21726
1st Coef(*100):	252059	130963
2nd Coef(*100):	270	387
H2O adj (mg/l*10k):	558	521

***** Printer Test End *****

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Print Test

Form 106-I8000

80-004189
6/7/11 YK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.13.00 06/09
01/06/2010 09:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:49
02 Std. Sol.	0.020	09:50
03 Room Air	0.000	09:51
04 Std. Sol.	0.020	09:52
05 Room Air	0.000	09:52
06 Std. Sol.	0.019	09:53
07 Room Air	0.000	09:53

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.019 AC

Form 106-18000

80-004189
6/9/10 YK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.13.00 06/09
01/06/2010 10:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:38
02 Std. Sol.	0.283	10:38
03 Room Air	0.000	10:39
04 Std. Sol.	0.284	10:40
05 Room Air	0.000	10:40
06 Std. Sol.	0.284	10:41
07 Room Air	0.000	10:42

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test

0.285AC

Form 106-I8000

80-004189
6/21/10

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.13.00 06/09
01/06/2010 13:20

Test	AC	Time
01 Room Air	0.000	13:21
02 *Subject Test	INT*	13:21
03 Room Air	0.000	13:22

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 659358
Cyl No = 34
Expiration Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Interferent test

Form 106-I8000

80-004189
67/1 FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.13.00 06/09
01/06/2010 14:45

Test	AC	Time
01 Diagnostic	OK	14:45
02 Room Air	RFI*	14:46
03 Room Air	0.000	14:46

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 659358
Cyl No = 34
Expiration Date = 09/03/2011
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

80-004189
6/21/11

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.13.00 06/09
01/06/2010 09:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:03
02 Std. Gas	0.079	09:03
03 Room Air	0.000	09:04
04 Std. Gas	0.079	09:04
05 Room Air	0.000	09:05
06 Std. Gas	0.079	09:05
07 Room Air	0.000	09:06

Lot No = 659358
Cyl No = 34
Exp Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004189
6/9/11 JAK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.13.00 06/09
01/06/2010 14:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:54
02 Std. Gas	0.079	14:54
03 Room Air	0.000	14:55
04 Std. Gas	0.079	14:55
05 Room Air	0.000	14:55
06 Std. Gas	0.079	14:56
07 Room Air	0.000	14:56

Lot No = 659358

Cyl No = 34

Exp Date = 09/03/2011

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004189

7/21/11

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.13.00 06/09
01/06/2010 14:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:59
02 Std. Gas	0.079	14:59
03 Room Air	0.000	14:59
04 Std. Gas	0.079	15:00
05 Room Air	0.000	15:00
06 Std. Gas	0.079	15:01
07 Room Air	0.000	15:01

Lot No = 659358

Cyl No = 34

Exp Date = 09/03/2011

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004189
7/21 KH