



OFFICE OF ATTORNEY GENERAL
Crime Laboratory Division
2641 East Main Avenue
P.O. Box 937
Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Billy Vannatta

Location: Mohl

Intoxilyzer Serial No.: 80-004198

Lot Number: 659358

Cylinder Number: 27

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.678 %

Test 2: 0.679 %

Test 3: 0.679 %

(ATTACH TEST RECORD)

Billy Vannatta 1-6-10
Chemical Test Operator Signature

Date Cylinder Changed

Deb Kashur
Reviewed By

08 Jan 2010
Date

Form 105-G
Toxicology Laboratory
Crime Laboratory Division

[Type text]

80-004198
P.142
VAT


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = MOHL 8164.13.00 06/09
01/06/2010 14:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:37
02 Std. Gas	0.078	14:37
03 Room Air	0.000	14:38
04 Std. Gas	0.079	14:38
05 Room Air	0.000	14:39
06 Std. Gas	0.079	14:39
07 Room Air	0.000	14:40

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 039909



Operator Signature
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-00478
P. 2 of 2 JFM