



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Matthew Anderson</u>		
Location: <u>Richland County Jail</u>		
Intoxilyzer® Serial Number: <u>80-004206</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>52</u>	Expiration Date: <u>9-30-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.083</u> %		
Test 2: <u>0.084</u> %		
Test 3: <u>0.084</u> %		
(ATTACH TEST RECORD)		

[Signature]
Chemical Test Operator Signature

1-06-11
Date

Deb Kashur
Reviewed By

10 Jan 2011
Date

80-004206
P. 110 of 111
SFN

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = WAHP 8164.13.00 06/09
01/06/2011 17:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:37
02 Std. Gas	0.083	17:37
03 Room Air	0.000	17:38
04 Std. Gas	0.084	17:38
05 Room Air	0.000	17:38
06 Std. Gas	0.084	17:39
07 Room Air	0.000	17:39

Lot No = 659358
Cyl No = 52
Exp Date = 09/03/2011
County = 39 Oper No. = 020401



Operator Signature
MATTHEW W. ANDERSON

Remarks:

Form 106-I8000

80-004206
P. 111 of 111
JEM