



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Haugen, Daniel</i> <i>NOHP</i>		
Location <i>Dickinson PD</i>	Intoxilyzer® Serial Number <i>90-002666</i>	
Gas Lot Number <i>23411090A1</i>	Gas Cylinder Number <i>21</i>	Gas Expiration Date <i>10/01/2013</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0. 080 AC
 - B. 0. 090 AC
 - C. 0. 081 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Daniel Haugen</i>	Date <i>01-06-2012</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kasher</i>	Date <i>03 Apr 2012</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = DKPD 8164.13.00 06/09
01/06/2012 10:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:11
02 Std. Gas	0.080	10:12
03 Room Air	0.000	10:12
04 Std. Gas	0.080	10:13
05 Room Air	0.000	10:13
06 Std. Gas	0.081	10:14
07 Room Air	0.000	10:14

Lot No = 23411080A1
Cyl No = 21
Exp Date = 10/01/2013
County = 45 Oper No. = 010105


Operator Signature
DANIEL J. HAUGEN

Remarks: *ACA for new cylinder.*