



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>RAY Weber</u>		
Location: <u>TRAIL CO. S.O.</u>		
Intoxilyzer® Serial Number: <u>80-004200</u>		
Gas Lot Number: <u>154100807A</u>	Gas Cylinder Number: <u>#15</u>	Expiration Date: <u>07/01/12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>078</u> % Test 2: 0. <u>077</u> % Test 3: 0. <u>077</u> %		
(ATTACH TEST RECORD)		

Ray Weber
 Chemical Test Operator Signature

1-8-2011
 Date

Deb Kashner
 Reviewed By

12 Jan 2011
 Date

80-004200
 P. 97 of 98
 VM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = HILL 8164.13.00 06/09
01/08/2011 07:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:56
02 Std. Gas	0.078	07:56
03 Room Air	0.000	07:57
04 Std. Gas	0.077	07:57
05 Room Air	0.000	07:58
06 Std. Gas	0.077	07:58
07 Room Air	0.000	07:58

Lot No = 15410080A1
Cyl No = 15
Exp Date = 07/01/2012
County = 49 Oper No. = 058405



Operator Signature
RAY G. WEBER

Remarks:

Form 106-I8000

80-004200
P. 98 of 98
JFM