

BREATH ALCOHOL INSTRUMENT CHAIN OF CUSTODY WORKSHEET

Instrument: 8000 Serial Number: 80-002666

Transported Out of the Crime Laboratory:

Agency Being Assigned Instrument: _____

Name/Agency of Person Transporting Instrument:

(Print Name) _____ (Print Agency)

Signature of Person Receiving Instrument: _____

Date Receiving Instrument: ____/____/____ From: _____

Or Shipped Via:
USPS ____ UPS ____ FedEx ____ Other _____

Date Sending Instrument: ____/____/____ Air Bill No. _____

Complete Air Bill (Identify Instrument/Serial Number) and attach a copy to this worksheet.

Transported to the Crime Laboratory:

Agency Returning Instrument: CMI

Name/Agency of Person Returning Instrument:

(Print Name) _____ (Print Agency)

Signature of Person Returning Instrument: _____

Date Returning Instrument: ____/____/____ Received By: Deb Kashur

Or Received Via:
USPS ____ UPS ____ FedEx Other _____

Date Receiving Instrument: 09 Jan 2013 Air Bill No. _____

Attach shipping paperwork (if any) to this worksheet.

Instrument Transfer Information Entered Into the Equipment Database:

Date of Action 22 Jan 2013 By: Deb Kashur



SERVICE WORK ORDER

DATE 11-26-12
 RCVD VIA FedEx

316 E. 9th St., Owensboro KY 42303
 Phone 866-835-0690
 Fax: 270-685-6268

330575

BILL TO:

ATTN:

CITY STATE ZIP

CUSTOMER NO.

PHONE

SHIP TO: ND, State of

ATTN: Crime Lab Div
2644 E. MAIN Ave.

CITY STATE ZIP
Bismarck ND 58501

PHONE 58501

FAX

MODEL # 8000 SERIAL # 80-002666 BILL CODE

EXTRA PARTS RCVD Gas Shaker

DESCRIPTION OF PROBLEM System Error 18 Ambient fails

Infrared / Fuel Cell
 PRELIMINARY TEST INFORMATION

BATTERY-EXT 12 V; INT 2928 V

System Error 18

VERIFIED PROBLEM? Yes

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	021307	Breath hose
1	013010	Battery
4	320384	Crimps

WORK PERFORMED:

Time/Date wrong - reset
Battery discharges rapidly
Breath hose/Cell vacuum leak
Cal/Sim vacuum OK
Pump Run OK
Snap top OK
RHS Flow Sensor OK
Keyboard retainer OK
RFI adjusted to .650V - Varnished
RTC OK

Miscellaneous
 PRELIMINARY TEST INFORMATION

WORK PERFORMED:
PST

REPAIR	CAL. ADJ.	CALIBRATION	FINAL
<u>JMT</u>	<u>DR</u>	<u>DR</u>	<u>DR</u>
1/2 HOURS	HOURS	HOURS	HOURS
<u>8</u>	<u>1</u>	<u>.5</u>	<u>.5</u>
DATE	DATE	DATE	DATE
<u>1-03-13</u>	<u>11-23-12</u>	<u>11-23-12</u>	<u>11-23-12</u>

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS