



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Phil Wolf</u>		
Location: <u>Cavalier, ND</u>		
Intoxilyzer® Serial Number: <u>80-002669</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>9</u>	Expiration Date: <u>4/3/11</u>
<input type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.077</u> %		
Test 2: <u>0.077</u> %		
Test 3: <u>0.077</u> %		
(ATTACH TEST RECORD)		

P. Wolf
Chemical Test Operator Signature

1/5/11
Date

Deb Kashur
Reviewed By

10 Jan 2011
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = CAVA 8164.13.00 06/09
01/05/2011 03:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	03:52
02 Std. Gas	0.077	03:52
03 Room Air	0.000	03:53
04 Std. Gas	0.077	03:53
05 Room Air	0.000	03:54
06 Std. Gas	0.077	03:54
07 Room Air	0.000	03:55

Lot No = 659357
Cyl No = 9
Exp Date = 09/03/2011
County = 34 Oper No. = 120547

P. Wolf

Operator Signature
PHIL WOLF

Remarks:

Form 106-I8000

80-002669
P. 103 of 103
jam