

Certificate of Calibration

Intoxilyzer® 8000

*This is to certify that the calibration of **Intoxilyzer**® serial number 80-004198, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.*

Date Jan 10, 2012 Signed Debi Rohrer

Technician

EMI INC.

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650519 CMI 7/9/09



SERVICE WORK ORDER

DATE	12-2-11
RCVD VIA	F.FOX

316 E. 9th St., Owensboro KY 42303
 Phone 866-835-0690
 Fax: 270-685-6268

3279164

BILL TO:		
ATTN:		
CITY	STATE	ZIP
CUSTOMER NO.	585013	
PHONE		

SHIP TO:		
N. D. Crime Lab		
ATTN: Deb Kashner		
2641 E. Main Ave.		
CITY	STATE	ZIP
Hickory	ND	58501
PHONE	701-328-1459	
FAX	701-328-1485	

MODEL #	8000	SERIAL #	SO-004198	BILL CODE	FLW
---------	------	----------	-----------	-----------	-----

EXTRA PARTS RCVD
 BT, gas box, battery

DESCRIPTION OF PROBLEM
 Flood sensor problem in the field.

Infrared / Fuel Cell
 PRELIMINARY TEST INFORMATION

Could not verify +
 Called and spoke w/ Deb Kashner

BATTERY- EXT 13.4 V INT 3.0 V
 VERIFIED PROBLEM? NO

WORK PERFORMED:

Replaced pressure sensor.
 Cleaned cell, inspected parts.
 Calibrated & fuel tested.
 Replaced external battery.

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	210120	Pressure sensor
1	013010	12v NiMH Battery

Miscellaneous
 PRELIMINARY TEST INFORMATION

ID# 04567

WORK PERFORMED:

REPAIR	CAL. ADJ.	CALIBRATION	FINAL
ML	DR	DR	DR
½ HOURS	HOURS	HOURS	HOURS
6	1	0	0.5
DATE	DATE	DATE	DATE
Jan 9 2012	Jan 13 2012	Jan 13 2012	Jan 13 2012

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS