

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-003067

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5144 Lot No. — Rep AC 0.000 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5187 Lot No. — Rep AC 0.285 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR5114 Lot No. — Rep AC 0.100 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 659358 Cyl No. 34 Exp. Date 03 Sept 2010

Test 1 0.079 Test 4 0.080 Test 7 0.079

Test 2 0.080 Test 5 0.079 Test 8 0.080

Test 3 0.080 Test 6 0.080 Test 9 0.080

Average 0.080

L. Remarks/Maintenance record: Software upgrade 8/6/13. Annual inspection.

12 Jan 2010
Date

Deb Kashur
Field Inspector's Signature

14 Jan 2010
Date

Mary Pearson
Reviewed by

80-003067
6/3/11 JAM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.13.00 06/09
01/11/2010 14:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:58
02 Std. Sol.	0.019	14:59
03 Room Air	0.000	14:59
04 Std. Sol.	0.018	15:00
05 Room Air	0.000	15:00
06 Std. Sol.	0.019	15:01
07 Room Air	0.000	15:01

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.019 AC

Form 106-I8000

80-003067
6/5/11 JKL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.13.00 06/09
01/11/2010 14:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:29
02 Std. Sol.	0.285	14:30
03 Room Air	0.000	14:31
04 Std. Sol.	0.284	14:31
05 Room Air	0.000	14:32
06 Std. Sol.	0.284	14:33
07 Room Air	0.000	14:33

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.285 AC

Form 106-I8000

80-003067
06/11/10

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.13.00 06/09
01/11/2010 16:19

Test	AC	Time
01 Room Air	0.000	16:19
02 *Subject Test	INT*	16:20
03 Room Air	0.000	16:20

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 659358
Cyl No = 34
Expiration Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Interferent check

Form 106-I8000

80-003067
6/21/11

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.13.00 06/09
01/12/2010 09:41

Test	AC	Time
01 Diagnostic	OK	09:42
02 Room Air	RFI*	09:42
03 Room Air	0.000	09:42

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 659358
Cyl No = 34
Expiration Date = 09/03/2011
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

80-003067
6/21/11 JH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.13.00 06/09
01/12/2010 09:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:45
02 Std. Gas	0.079	09:45
03 Room Air	0.000	09:45
04 Std. Gas	0.080	09:46
05 Room Air	0.000	09:46
06 Std. Gas	0.080	09:47
07 Room Air	0.000	09:47

Lot No = 659358
Cyl No = 34
Exp Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check

Form 106-I8000

80-003067
6/9/11 *AK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.13.00 06/09
01/12/2010 09:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:52
02 Std. Gas	0.080	09:52
03 Room Air	0.000	09:53
04 Std. Gas	0.079	09:53
05 Room Air	0.000	09:54
06 Std. Gas	0.080	09:54
07 Room Air	0.000	09:54

Lot No = 659358
Cyl No = 34
Exp Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks: *Calibration check*

Form 106-I8000

*80-003067
7/21 FH*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.13.00 06/09
01/12/2010 09:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:56
02 Std. Gas	0.079	09:56
03 Room Air	0.000	09:56
04 Std. Gas	0.080	09:57
05 Room Air	0.000	09:57
06 Std. Gas	0.080	09:58
07 Room Air	0.000	09:58

Lot No = 659358
Cyl No = 34
Exp Date = 09/03/2011
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Calibration check

Form 106-18000

80-003067
7/21/11 PK