



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59262 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Mark W. McDonald</u>		
Location: <u>Valley City Police Dept.</u>		
Intoxilyzer® Serial Number: <u>80-004201</u>		
Gas Lot Number: <u>659356</u>	Gas Cylinder Number: <u>24</u>	Expiration Date: <u>09/03/2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.079</u> %		
Test 2: <u>0.079</u> %		
Test 3: <u>0.078</u> %		
(ATTACH TEST RECORD)		

[Signature]
Chemical Test Operator Signature

01/13/2011
Date

Deb Kashner
Reviewed By

24 Jan 2011
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = VALL 8164.13.00 06/09
01/13/2011 16:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:05
02 Std. Gas	0.079	16:06
03 Room Air	0.000	16:06
04 Std. Gas	0.079	16:06
05 Room Air	0.000	16:07
06 Std. Gas	0.078	16:07
07 Room Air	0.000	16:08

Lot No = 659358
Cyl No = 24
Exp Date = 09/03/2011
County = 02 Oper No. = 119502



Operator Signature
MARK W. MCDONALD

Remarks:

Form 106-I8000

80-004201
P. 97 of 97
1/13