



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (03-2014)

Serial Number 80-004952	Instrument Location HETT
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input checked="" type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): NA
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>J Collins</i>	Date <i>01/13/2015</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>21 Jan 15</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = HETT 8164.13.00 06/09
01/13/2015 08:52

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 33913080A4
Standard Cyl #? 13
Standard Expiration? 01/01/2016
Oper No? 020533

Flow Cal. Date: 01/19/2010
Slope 694
Intercept -638616

IR Calibration Date: 01/19/2010
 3um 9um

0th Coef(*100): -32832 -34603
1st Coef(*100): 262566 131011
2nd Coef(*100): 1418 975
H2O adj(mg/l*10k): 686 550

***** Printer Test End *****



Operator Signature
TRAVIS COLLINS

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = HETT 8164.13.00 06/09
01/13/2015 08:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:53
02 Std. Gas	0.080	08:53
03 Room Air	0.000	08:54
04 Std. Gas	0.080	08:54
05 Room Air	0.000	08:55
06 Std. Gas	0.080	08:55
07 Room Air	0.000	08:56

Lot No = 33913080A4
Cyl No = 13
Exp Date = 01/01/2016
County = 01 Oper No. = 020533



Operator Signature
TRAVIS COLLINS

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = HETT 8164.13.00 06/09
01/13/2015 08:56

Test	AC	Time
01 Room Air	0.000	08:58
02 *Subject Test	RFI*	08:58
03 Room Air	0.000	08:58

*Invalid Test
Inhibited - RFI

Sub Name = TEST, LOCT A
Sub DOB = 01/01/2012
Sub Sex = Male Weight = 120
Test = OTH Cit = 0000000
Dr. Lic. = ND/1234567
Lot No = 33913080A4
Cyl No = 13
Expiration Date = 01/01/2016
County = 01 Oper No. = 020533



Operator Signature
TRAVIS COLLINS

Remarks:

Form 106-I8000