

INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006499

1) GENERAL Setup and Checks:

- a) Power On
 - 5 min. countdown during initial warm-up
 - 1 min. countdown to begin test after standby mode
- b) Display Intensity (Ctrl + ↑ or ↓)
- c) Password Security:
 - Level 1 (Operator) Level 2 (Field Inspector) Level 3 (Lab Personnel)
- d) Printer Setup (Level 3,S,P)
- e) Forms Load (CMI, Inc., CoBrA Software)

2) LEVEL 1 Setup and Checks:

- a) Internal Printer Test and Paper Feed (Level 1,P). Attach record.
- b) External Printer Test (Level 1,P). Attach record.
- c) Barcode Reader (Level 1,B):
 - Operator ID ND License R key: Refusal. Attach record.

3) LEVEL 2 Setup and Checks:

- a) Continuous Air Blank (Level 2,A)
- b) Diagnostic (Level 2,D). Attach record.
- c) Prelim Data Entry (Level 2,E)
- d) Ser Number/Version Info (Level 2,V):
 - Inst Ser#: 80-006499 DSP Ver: 1362.14
 - Core Ver: 8164.13 CPLD Ver: 7

4) LEVEL 3 Setup and Checks:

- a) DVM Monitor (Level 3,D,D): 3: 12301 s: 2 C: 47.06
9: 13251 s: 1 B: 45 0
(RFI indicator with keyed radio): 1
- b) Function (F) Keys: : F5 – Tone F9 - Fan
 F6 – Green LED F10 – Pump
 F7 – Red LED F11 – Solenoid
- c) Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 946 CF: 1.07
- d) Flow Monitor (Level 3,D,F): Flow Sensor, R: 247 F: 0.000 L/s

- e) Temperature Monitor (Level 3,D,T): C: 47.0 A: 49.5 I: 0.0
B: 45.1 V: 0.0
- f) Tank Monitor (Level 3,D,G): P (display): 961 psi P (gauge): ≈970 psi
- g) General Setup (Level 3,S,G)
- h) Configure Start Test (Level 3,S,T)
- i) Network (Level 3,S,N)

5) TESTS (attach all test records):

Configure wet bath simulator standards before running tests (Level 1,S):

- a) Zero AC (Level 1,C):
Sim ser #: DR3375 Lot #: (N/A) (Milli Q) Exp date: N/A
Known value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
- b) Low AC (Level 1,C):
Sim ser #: DR3374 Lot #: 14240 Exp date: 10/20/16
Known value: 0.020 AC
Results: 0.019 AC 0.019 AC 0.019 AC
- c) High AC (Level 1,C):
Sim ser #: DR3841 Lot #: 201408c Exp date: 08/12/16
Known value: 0.300
Results: 0.295 AC 0.296 AC 0.296 AC

Configure dry gas standard before running tests (Level 1,S):

- d) Ethanol + 0.05% Acetone (Level 1,B):
Sim ser #: DR5144 Lot #: ICS4 Exp date: ^{DS} ~~09/17~~ ^{DS} 09/09/17
Known value: 0.10 AC + 0.05% acetone
 Display reads "Interferent Detect"
- e) Acetone (Level 1,B):
Sim ser #: DR3455 Lot #: Acc1 Exp date: 09/29/17
Known value: 0.05% acetone
 Display reads "Interferent Detect"
- f) Methanol (Level 1,B):
Sim ser #: DR5189 Lot #: Met2 Exp date: 09/29/17
Known value: 0.10% Methanol
 Display reads "Interferent Detect"
- g) Isopropanol (Level 1,B):
Sim ser #: DR3454 Lot #: IPa2 Exp date: 09/29/17
Known value: 0.10% IPA
 Display reads "Interferent Detect"

- h) Ambient Fail (Level 1,B): Display and test record read correct.
- i) Invalid Sample (Level 1,B): Display and test record read correct.
- j) Sequence Aborted (Level 1,B): Display and test record read correct.
- k) ABA (Level 1,B): Test record reads correct.
- l) Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- m) RFI Detected (B,C, or CMS mode): Display and test record read correct.
- n) Tank Pressure Below Minimum (CMS mode): Display reads correct.
- o) Deficient Samples (CMS mode): Display and test record read correct.
- p) Improper Sample (CMS mode): Display and test record read correct.
- q) Difference Too Great (CMS mode): Display and test record read correct.
- r) CMS (CMS mode): Test record reads correct.
- s) Stability Test (Level 3,T,S):

Ave: 0.0814 Std Dev: 0.0065 Rel Std Dev: 0.61%

Maintenance/Repairs/Notes: _____

Instrument is acceptable to be used in the field.

D. b. Shearer
Field Inspector's Signature

13 Jan 2016
Date

Charles E. Elder
Reviewed by

14 JAN 16
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMi, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 09:54

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

24 characters per line
abcdefghijklmnopqrstuwx
ABCDEFGHIJKLMNQRSTUWX

32 characters per line
abcdefghijklmnopqrstuwx123456
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuwx1234567890abcdef
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup
Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Menfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 87653
Standard Cyl #? 12357
Standard Expiration? 11/11/2020
Oper No? 77777

Flow Cal. Date: 08/18/2015
Slope 678
Intercept -685002

IR Calibration Date: 08/18/2015
3um 9um

0th Coef(*100):	-22123	-16014
1st Coef(*100):	264933	138505
2nd Coef(*100):	3281	1621
H2O adj(mg/l)*10K):	704	456

***** Test End *****

Deb Shanauer
Operator Signature
DEB SHANAUER

Remarks:

Print test

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 11:08

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNQPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNQPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 29614080A2
Standard Cyl #? 25
Standard Expiration? 12/05/2016
Oper No? 777777

Flow Cal. Date: 08/18/2015
Slope 678
Intercept -685002

IR Calibration Date: 08/18/2015
 3um 9um

	3um	9um
0th Coef(*100):	-22123	-16014
1st Coef(*100):	264933	138505
2nd Coef(*100):	3281	1621
H2O adj(mg/l*10k):	704	456

***** Printer Test End *****



Operator Signature
DEB SHANAVER

Remarks:

Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 11:08

Test	AC	Time
01 Room Air	0.000	11:09
02 *Subject Test	REF*	11:09
03 Room Air	0.000	11:10

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:



Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 11:11

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Diagnostic test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 11:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:24
02 Std. Sol.	0.000	11:25
03 Room Air	0.000	11:25
04 Std. Sol.	0.000	11:26
05 Room Air	0.000	11:26
06 Std. Sol.	0.000	11:27
07 Room Air	0.000	11:28

08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = MILLIQ

County = 08

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Zero alcohol 0.000AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 14:00

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:01
02 Std. Sol.	0.019	14:02
03 Room Air	0.000	14:03
04 Std. Sol.	0.019	14:04
05 Room Air	0.000	14:04
06 Std. Sol.	0.019	14:05
07 Room Air	0.000	14:05

08 Sim Temp = 34.0°C

Simul Ser No = DR3374

Std Sol No = 14240

County = 08

Oper No. = 777777

Deb Shanaver

Operator Signature

DEB SHANAVER

Remarks:

Low AC 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 14:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:09
02 Std. Sol.	0.295	14:10
03 Room Air	0.000	14:11
04 Std. Sol.	0.296	14:12
05 Room Air	0.000	14:12
06 Std. Sol.	0.296	14:13
07 Room Air	0.000	14:14

08 Sim Temp = 34.0°C

Simul Ser No = DR3841

Std Sol No = 201408C

County = 08

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

High AC 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 14:16

Test	AC	Time
01 Room Air	0.000	14:17
02 *Subject Test	INT*	14:17
03 Room Air	0.000	14:18

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.10AC + 0.05% acetone

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 14:18

Test	AC	Time
01 Room Air	0.000	14:19
02 *Subject Test	INT*	14:19
03 Room Air	0.000	14:20

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 77777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 14:21

Test	AC	Time
01 Room Air	0.000	14:21
02 *Subject Test	INT*	14:22
03 Room Air	0.000	14:22

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.10% methanol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 14:25

Test	AC	Time
01 Room Air	0.000	14:26
02 *Subject Test	INT*	14:26
03 Room Air	0.000	14:27

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

DS Interferent detected
~~Ambient fail~~ 0.1090 isopropanol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 14:29

Test	AC	Time
01 Room Air	AMB*	14:29
02 Room Air	0.000	14:30

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:



Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 14:36

Test	AC	Time
01 Room Air	0.000	14:37
02 Invalid Sample	X.XXX	14:37
03 Room Air	0.000	14:38

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Invalid sample

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 14:30

Test	AC	Time
01 Room Air	0.000	14:31
02 *Subject Test	ABT*	14:31
03 Room Air	0.000	14:31

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Sequence aborted

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 14:42

Test	AC	Time
01 Room Air	0.000	14:43
02 Subject Test 1	0.000	14:43
03 Room Air	0.000	14:43
04 Reported AC	0.000	14:43

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

ABA test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 16:08

Test	AC	Time
01 Diagnostic	OK	16:09
02 Room Air	0.000	16:09
03 Subject Test 1	0.000	16:09
04 Room Air	0.000	16:12
05 Std. Gas	0.080*	16:13
06 Room Air	0.000	16:13

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Calibration check out of tolerance

Form 106-I8000

0.070 was entered as the gas AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 14:40

Test	AC	Time
01 Diagnostic	OK	14:41
02 Room Air	RFI*	14:41
03 Room Air	0.000	14:41

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

RFI test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 16:23

Test	AC	Time
01 Diagnostic	OK	16:24
02 Room Air	0.000	16:24
03 *Subject Test	0.000*	16:27
04 Room Air	0.000	16:30
05 Std. Gas	0.079	16:31
06 Room Air	0.000	16:32
07 *Subject Test	0.000*	16:35
08 Room Air	0.000	16:37


No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777


20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:



Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 16:38

Test	AC	Time
01 Diagnostic	OK	16:39
02 Room Air	IPS*	16:39
03 Room Air	0.000	16:39

*Improper Sample

Sub Name = , THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Improper sample

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 16:40

Test	AC	Time
01 Diagnostic	OK	16:41
02 Room Air	0.000	16:41
03 Subject Test 1	0.000*	16:41
04 Room Air	0.000	16:43
05 Std. Gas	0.079	16:45
06 Room Air	0.000	16:46
07 Subject Test 2	0.043*	16:46
08 Room Air	0.000	16:47

*Difference Too Great


Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:



Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 16:48


Test	AC	Time
01 Diagnostic	OK	16:49
02 Room Air	0.000	16:49
03 Subject Test 1	0.000	16:49
04 Room Air	0.000	16:51
05 Std. Gas	0.081	16:53
06 Room Air	0.000	16:54
07 Subject Test 2	0.000	16:54
08 Room Air	0.000	16:55
09 Reported AC	0.000	16:49

Difference OK
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:

CMS test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.13.00 06/09
01/13/2016 16:56

Test	AC	Time
01 Room Air	0.000	16:57
02 Std. Gas	0.081	16:57
03 Room Air	0.000	16:58
04 Std. Gas	0.081	16:58
05 Room Air	0.000	16:59
06 Std. Gas	0.081	16:59
07 Room Air	0.000	16:59
08 Std. Gas	0.081	17:00
09 Room Air	0.000	17:00
10 Std. Gas	0.081	17:01
11 Room Air	0.000	17:01
12 Std. Gas	0.081	17:02
13 Room Air	0.000	17:02
14 Std. Gas	0.081	17:02
15 Room Air	0.000	17:03
16 Std. Gas	0.082	17:03
17 Room Air	0.000	17:04
18 Std. Gas	0.082	17:04
19 Room Air	0.000	17:05
20 Std. Gas	0.081	17:05
21 Room Air	0.000	17:06
22 Std. Gas	0.081	17:06
23 Room Air	0.000	17:07
24 Std. Gas	0.081	17:07
25 Room Air	0.000	17:07
26 Std. Gas	0.081	17:08
27 Room Air	0.000	17:08
28 Std. Gas	0.082	17:09
29 Room Air	0.000	17:09
30 Std. Gas	0.082	17:10
31 Room Air	0.000	17:10
32 Std. Gas	0.081	17:10
33 Room Air	0.000	17:11
34 Std. Gas	0.081	17:11
35 Room Air	0.000	17:12
36 Std. Gas	0.082	17:12
37 Room Air	0.000	17:13
38 Std. Gas	0.082	17:13
39 Room Air	0.000	17:14
40 Std. Gas	0.082	17:14
41 Room Air	0.000	17:15
42 Cal Check Stats:		
43 Average	0.0814	
44 Std Dev	0.0005	
45 Rel Std Dev	0.61%	

*Stability test
Dob Shanaver*

Operator Signature
DEB SHANAVER

Remarks:

lot: 29614680A2 cyl: 25 exp: 12/05/2016

Form 106-I8000

Stability test
Deb Shanaver