



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Aaron Geddes</i>		
Location <i>Dunn County Sheriff Office</i>	Intoxilyzer® Serial Number <i>86-005755</i>	
Gas Lot Number <i>19615080A5</i>	Gas Cylinder Number <i>29</i>	Gas Expiration Date <i>01/05/2017</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.083* AC
 - B. *0.083* AC
 - C. *0.084* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>01/14/16</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Shanaver</i>	Date <i>22 Jan 16</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.13.00 06/09
01/14/2016 16:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:28
02 Std. Gas	0.083	16:28
03 Room Air	0.000	16:29
04 Std. Gas	0.083	16:29
05 Room Air	0.000	16:30
06 Std. Gas	0.084	16:30
07 Room Air	0.000	16:31

Lot No = 19615080A5
Cyl No = 29
Exp Date = 09/05/2017
County = 13 Oper No. = 130401



Operator Signature
AARON GEDDES

Remarks:

Form 106-I8000