

**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**

Serial No.: 80-003061 Instrument Location: Williston LEC

Reason for Install/Repair: \_\_\_\_\_ Install After Lab Repair/Inspection  
X Other (Specify) New

Check When Done:

- X A. Surge Protector Installed/Properly Grounded.
- X B. Telephone Line Connected to Intoxilyzer®.
- X C. Breath Tube Heated.
- X D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- X E. Review/Enter Gas Setup (Level 1, Function S).
- X F. Print Test (Level 1, Function P). (Attach Test Record.)
- X G. ACA Test (Level 1, Function C). (Attach Test Record.)
- X H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: \_\_\_\_\_

01-15-10  
Date

Darryl W. Aberle  
Field Inspector's Signature

20 Jan 2016  
Date

Deb Kashner  
Reviewed By

80-003061  
p. 1 of 4  
FM




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = ~~FOX~~ WILL OK      8164.13.00 06/09  
01/15/2010      08:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:13
02 Std. Gas	0.078	08:14
03 Room Air	0.000	08:14
04 Std. Gas	0.079	08:15
05 Room Air	0.000	08:15
06 Std. Gas	0.078	08:16
07 Room Air	0.000	08:16

Lot No = 659357  
Cyl No = 45  
Exp Date = 09/03/2011  
County = 53      Oper No. = 010101

  
Operator Signature  
DARCY W. ABERLE

Remarks:

ACA Test 1

Form 106-I8000

80-003061  
P. 3 of 4  
JFM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-003061  
Location = ~~TOXEL WILL~~ <sup>DK</sup>                      8164.13.00 06/09  
01/15/2010    08:17

Test	AC	Time
01 Diagnostic	OK	08:18
02 Room Air	RFI*	08:19
03 Room Air	0.000	08:19

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI, TEST T  
Sub DOB = 07/04/1980  
Sub Sex = Male                                      Weight = 200  
Test = OTH    Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 659357  
Cyl No = 45  
Expiration Date = 09/03/2011  
County = 53                                      Oper No. = 010101

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
DARCY W. ABERLE

Remarks:

Form 106-I8000

80-003061  
P. 4/1/11