

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Darcy W Aberle

Location: Williston LEC

Intoxilyzer Serial No.: 80-003061

Lot Number: 659357

Cylinder Number: 45

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.080%

Test 2: 0.079%

Test 3: 0.078%

(ATTACH TEST RECORD)

Darcy W Aberle
Chemical Test Operator Signature

01-15-10
Date

Deb Kashner
Reviewed By

20 Jan 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003061
Location = ~~FOX~~WILL ^{DK} 8164.13.00 06/09
01/15/2010 08:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:20
02 Std. Gas	0.080	08:20
03 Room Air	0.000	08:21
04 Std. Gas	0.079	08:21
05 Room Air	0.000	08:22
06 Std. Gas	0.078	08:22
07 Room Air	0.000	08:23

Lot No = 659357
Cyl No = 45
Exp Date = 09/03/2011
County = 53 Oper No. = 010101



Operator Signature
DARCY W. ABERLE

Remarks:

Form 106-I8000

80-003061
P. 2 of 2 *FW*